POLICE DEPARTMENT
**LAKE HAVASU CITY**
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Karl C. E. Dreller is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are a full, true and correct copy of the original.

K.C. 121
SQAP SIGNATURE 01/12/2009

STATE OF ARIZONA )
) ss
COUNTY OF MOHAVE )

Subscribed and sworn to before me this 12TH day of JANUARY, 2009 by KC DRELLER.

Gloria Thomas
NOTARY PUBLIC

1-12-09
DATE

Commission expires: 8-31-10

Notary Public State of Arizona
Mohave County
Gloria Thomas
My Commission Expires
08/31/2010

Area Code 928

Administration ..... 855-4884
Business Office ..... 855-1171
Investigations ..... 855-5775
Patrol ..... 855-0515

Fax #'s
Administration ..... 680-5430
Business Office ..... 680-5431
Investigations/Patrol ..... 680-5432

Emergency ..... 9-1-1
Non-Emergency Dispatch/ ..... 855-4111
*TDD* 855-4114
Havasu Silent Witness ..... 854-TIPS

E-mail: police@ci.lake-havasu-city.az.us

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K.C. Dreller #121 AGENCY Lake Havasu City Police Dept.

DATE 01/22/2009 TIME 1044

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD Main

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
   OR
   Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.099 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K.C. Dreller 121

K.C. Dreller 121
Detective

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number: 80-001029
01/12/2009 10:44:59

---

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:45:26 |
| Cal Check | 0.099 | 10:45:45 |
| Air Blank | 0.000 | 10:46:14 |

K. C. Dreller 121

Operator's Signature

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K.C. DAELLER #121 AGENCY LAKE HAVASU CITY POLICE DEPT

DATE 011209 TIME 1047

INTOXILYZER SERIAL # 80·001029 LOCATION LHCPD MAIN

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
1. Clock time check.
2. Date check.

OPERATIONAL TESTS
1. Alcohol-free subject test result 0. 000 AC.
2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE K.C. 121
K.C. DAELLER #121
DETECTIVE

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
01/12/2009 10:47:30

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,FREE
DOB: 01/12/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:48:39 |
| Diagnostic Test | Pass | 10:49:06 |
| Air Blank | 0.000 | 10:49:34 |
| 0.100 Cal Check | 0.100 | 10:49:53 |
| Air Blank | 0.000 | 10:50:22 |
| Subject Test | 0.000 | 10:50:47 |
| Air Blank | 0.000 | 10:51:15 |
| Five Minute Wait | ABT* | |
| Air Blank | 0.000 | 10:52:04 |

*Sequence Aborted

**Not a Successfully Completed Test Sequence**

121

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
01/12/2009 10:47:30

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,FREE
DOB: 01/12/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:48:39 |
| Diagnostic Test | Pass | 10:49:06 |
| Air Blank | 0.000 | 10:49:34 |
| 0.100 Cal Check | 0.100 | 10:49:53 |
| Air Blank | 0.000 | 10:50:22 |
| Subject Test | 0.000 | 10:50:47 |
| Air Blank | 0.000 | 10:51:15 |
| Five Minute Wait | ABT* | |
| Air Blank | 0.000 | 10:52:04 |

*Sequence Aborted

**Not a Successfully Completed Test Sequence**

KD 121

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
01/12/2009 10:47:30

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,FREE
DOB: 01/12/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:48:39 |
| Diagnostic Test | Pass | 10:49:06 |
| Air Blank | 0.000 | 10:49:34 |
| 0.100 Cal Check | 0.100 | 10:49:53 |
| Air Blank | 0.000 | 10:50:22 |
| Subject Test | 0.000 | 10:50:47 |
| Air Blank | 0.000 | 10:51:15 |
| Five Minute Wait | ABT* | |
| Air Blank | 0.000 | 10:52:04 |

*Sequence Aborted

**Not a Successfully Completed Test Sequence**

121

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
01/12/2009 10:52:29

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,BAWT
DOB: 01/12/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | IPS* | 10:53:23 |
| Air Blank | 0.000 | 10:53:51 |

*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
01/12/2009 10:52:29

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,BAWT
DOB: 01/12/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | IPS* | 10:53:23 |
| Air Blank | 0.000 | 10:53:51 |

*Improper Sample

**Not a Successfully Completed Test Sequence**

121

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
01/12/2009 10:52:29

---

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,BAWT
DOB: 01/12/2009
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | IPS* | 10:53:23 |
| Air Blank | 0.000 | 10:53:51 |

*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
01/12/2009 10:54:14

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,RFI
DOB: 01/12/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | RFI* | 10:55:15 |
| Air Blank | 0.000 | 10:55:44 |

*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
01/12/2009 10:54:14

---

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,RFI
DOB: 01/12/2009
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | RFI* | 10:55:15 |
| Air Blank | 0.000 | 10:55:44 |

*RFI Detect

**Not a Successfully Completed Test Sequence**

RCD/121

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
01/12/2009 10:54:14

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,RFI
DOB: 01/12/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | RFI* | 10:55:15 |
| Air Blank | 0.000 | 10:55:44 |

*RFI Detect

**Not a Successfully Completed Test Sequence** KD/121

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
01/12/2009 10:56:06

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,DEFSAMP
DOB: 01/12/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:57:12 |
| Diagnostic Test | Pass | 10:57:39 |
| Air Blank | 0.000 | 10:58:06 |
| 0.100 Cal Check | 0.100 | 10:58:25 |
| Air Blank | 0.000 | 10:58:55 |
| Subject Test | 0.436* | 11:02:03 |
| Air Blank | 0.000 | 11:02:36 |
| Air Blank | 0.000 | 11:03:04 |
| Subject Test | XXX** | 11:03:32 |
| Air Blank | 0.000 | 11:04:07 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
01/12/2009 10:56:06

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,DEFSAMP
DOB: 01/12/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:57:12 |
| Diagnostic Test | Pass | 10:57:39 |
| Air Blank | 0.000 | 10:58:06 |
| 0.100 Cal Check | 0.100 | 10:58:25 |
| Air Blank | 0.000 | 10:58:55 |
| Subject Test | 0.436* | 11:02:03 |
| Air Blank | 0.000 | 11:02:36 |
| Air Blank | 0.000 | 11:03:04 |
| Subject Test | XXX** | 11:03:32 |
| Air Blank | 0.000 | 11:04:07 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number : 80-001029
01/12/2009 10:56:06

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,DEFSAMP
DOB: 01/12/2009
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:57:12 |
| Diagnostic Test | Pass | 10:57:39 |
| Air Blank | 0.000 | 10:58:06 |
| 0.100 Cal Check | 0.100 | 10:58:25 |
| Air Blank | 0.000 | 10:58:55 |
| Subject Test | 0.436* | 11:02:03 |
| Air Blank | 0.000 | 11:02:36 |
| Air Blank | 0.000 | 11:03:04 |
| Subject Test | XXX** | 11:03:32 |
| Air Blank | 0.000 | 11:04:07 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

KD/121

**EXHIBIT G-2**
**THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)**

**ARIZONA DEPARTMENT OF PUBLIC SAFETY**

**STANDARD QUALITY ASSURANCE PROCEDURES**
**INTOXILYZER MODEL 8000**

**STANDARD CALIBRATION CHECK PROCEDURE**

QA SPECIALIST K.C. Dreller #121 AGENCY Lake Havasu City Police Dept.

DATE 01 12 2009 TIME 1104

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD Main

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

   OR

   Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.098 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K.C. Dreller #121

K.C. Dreller #121
Detective

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LAKE HAVASU PD
Serial Number: 80-001029
01/12/2009 11:04:40

---

Standard Lot#: 727501I
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:05:07 |
| Cal Check | 0.098 | 11:05:26 |
| Air Blank | 0.000 | 11:05:56 |

Operator's Signature