POLICE DEPARTMENT
**LAKE HAVASU CITY**
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

J Sondrol 187
Signature

STATE OF ARIZONA )
) ss
COUNTY OF MOHAVE )

Subscribed and sworn to before me this 4 day of February, 2013 by Carla Duarte

Carla Duarte
NOTARY PUBLIC

2-4-13
DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2013

Area Code 928

Administration ..... 855-4884
Business Office ..... 855-1171
Investigations ..... 855-5775
Patrol ..... 855-0515

Fax #'s
Administration ..... 680-5430
Business Office ..... 680-5431
Investigations/Patrol ..... 680-5432

Emergency ..... 9-1-1
Non-Emergency Dispatch/ ..... 855-4111
TDD 855-4114
Havasu Silent Witness ..... 854-TIPS

E-mail: police@lhcaz.gov

**EXHIBIT G-2**
**THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)**

**ARIZONA DEPARTMENT OF PUBLIC SAFETY**

**STANDARD QUALITY ASSURANCE PROCEDURES**
**INTOXILYZER MODEL 8000**

**STANDARD CALIBRATION CHECK PROCEDURE**

QA SPECIALIST J. SONDROL 187 AGENCY LAKE HAVASU CITY PD

DATE 1/28/13 TIME 2011

INTOXILYZER SERIAL # 80-001029 LOCATION LHC MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.098 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

*CHANGED STANDARD PRIOR

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
01/28/2013 20:11:53

---

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 20:12:23 |
| 0.100 Cal Check | 0.098 | 20:12:44 |
| Air Blank | 0.000 | 20:13:13 |

Preventative Maintenance
Performed -- Timer Reset

**EXHIBIT G-4**
**THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)**

**ARIZONA DEPARTMENT OF PUBLIC SAFETY**

**STANDARD QUALITY ASSURANCE PROCEDURES**
**INTOXILYZER MODEL 8000**

**STANDARD QUALITY ASSURANCE PROCEDURE**

QA SPECIALIST J. SONDROL 187 AGENCY LAKE HAVASU CITY PD

DATE 1/28/13 TIME 2015

INTOXILYZER SERIAL # 80-001029 LOCATION LNC MAIN JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE J Sondrol 187

DPS Form Exh G-4 (Rev 05-01)

gs/187

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
01/28/2013 20:15:30

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 20:16:32 |
| Diagnostic Test | Pass | 20:17:07 |
| Air Blank | 0.000 | 20:17:35 |
| 0.100 Cal Check | 0.099 | 20:17:55 |
| Air Blank | 0.000 | 20:18:25 |
| Subject Test | 0.000 | 20:18:43 |
| Air Blank | ABT* | 20:18:53 |
| Air Blank | IPS** | 20:19:01 |

*Sequence Aborted
**Improper Sample

**Not a Successfully Completed Test Sequence**

9/187

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
01/28/2013 20:15:30

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 20:16:32 |
| Diagnostic Test | Pass | 20:17:07 |
| Air Blank | 0.000 | 20:17:35 |
| 0.100 Cal Check | 0.099 | 20:17:55 |
| Air Blank | 0.000 | 20:18:25 |
| Subject Test | 0.000 | 20:18:43 |
| Air Blank | ABT* | 20:18:53 |
| Air Blank | IPS** | 20:19:01 |

*Sequence Aborted
**Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
01/28/2013 20:15:30

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 20:16:32 |
| Diagnostic Test | Pass | 20:17:07 |
| Air Blank | 0.000 | 20:17:35 |
| 0.100 Cal Check | 0.099 | 20:17:55 |
| Air Blank | 0.000 | 20:18:25 |
| Subject Test | 0.000 | 20:18:43 |
| Air Blank | ABT* | 20:18:53 |
| Air Blank | IPS** | 20:19:01 |

*Sequence Aborted
**Improper Sample

**Not a Successfully Completed Test Sequence**

7187

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
01/28/2013 20:20:53

---

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST
DOB: 10/10/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 20:21:53 |
| Diagnostic Test | Pass | 20:22:28 |
| Air Blank | 0.000 | 20:22:55 |
| 0.100 Cal Check | 0.099 | 20:23:16 |
| Air Blank | 0.000 | 20:23:45 |
| Subject Test | 0.057* | 20:26:53 |
| Air Blank | 0.000 | 20:27:25 |
| Air Blank | 0.000 | 20:27:52 |
| Subject Test | XXX** | 20:28:16 |
| Air Blank | RFI*** | 20:28:42 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

**Not a Successfully Completed Test Sequence**

4187

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
01/28/2013 20:20:53

---

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST
DOB: 10/10/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 20:21:53 |
| Diagnostic Test | Pass | 20:22:28 |
| Air Blank | 0.000 | 20:22:55 |
| 0.100 Cal Check | 0.099 | 20:23:16 |
| Air Blank | 0.000 | 20:23:45 |
| Subject Test | 0.057* | 20:26:53 |
| Air Blank | 0.000 | 20:27:25 |
| Air Blank | 0.000 | 20:27:52 |
| Subject Test | XXX** | 20:28:16 |
| Air Blank | RFI*** | 20:28:42 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
01/28/2013 20:20:53

---

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST
DOB: 10/10/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 20:21:53 |
| Diagnostic Test | Pass | 20:22:28 |
| Air Blank | 0.000 | 20:22:55 |
| 0.100 Cal Check | 0.099 | 20:23:16 |
| Air Blank | 0.000 | 20:23:45 |
| Subject Test | 0.057* | 20:26:53 |
| Air Blank | 0.000 | 20:27:25 |
| Air Blank | 0.000 | 20:27:52 |
| Subject Test | XXX** | 20:28:16 |
| Air Blank | RFI*** | 20:28:42 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. SONDROL 187 AGENCY LAKE HAVASU CITY PD

DATE 1/28/13 TIME 2031

INTOXILYZER SERIAL # 80-001029 LOCATION LHC MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.098 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

7101

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
01/28/2013 20:31:48

---

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 20:32:19 |
| 0.100 Cal Check | 0.098 | 20:32:40 |
| Air Blank | 0.000 | 20:33:10 |

Preventative Maintenance
Performed -- Timer Reset