POLICE DEPARTMENT
**LAKE HAVASU CITY**
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Karl C. E. Dreller is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are a full, true and correct copy of the original.

K.C.E. Dreller
SQAP SIGNATURE 02032011
80-001029
END OF STANDARD

STATE OF ARIZONA )
) ss
COUNTY OF MOHAVE )

Subscribed and sworn to before me this 7 day of FEBRUARY, 2011 by CARLA DUARTE.

Carla Duarte
NOTARY PUBLIC

2-7-11
DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2013

Area Code 928

Administration ..... 855-4884
Business Office ..... 855-1171
Investigations ..... 855-5775
Patrol ..... 855-0515

Fax #'s
Administration ..... 680-5430
Business Office ..... 680-5431
Investigations/Patrol ..... 680-5432

E-mail: police@ci.lake-havasu-city.az.us

Emergency ..... 9-1-1
Non-Emergency Dispatch/ ..... 855-4111
*TDD* 855-4114
Havasu Silent Witness ..... 854-TIPS

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R.C. DRELLER #121 AGENCY LAKE HAVASU CITY POLICE DEPT.

DATE 02 03 2011 TIME 0959

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD MAIN JAIL

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution. 100 AC.

   OR

   Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.101 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE R.C. Dreller #121

R.C. DRELLER #121
DETECTIVE

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 09:59:17

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 09:59:52 |
| 0.100 Cal Check | 0.101 | 10:00:13 |
| Air Blank | 0.000 | 10:00:43 |

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K.C. DRELLER AGENCY LAKE HAVASU CITY POLICE DEPT

DATE 02032011 TIME 1009

INTOXILYZER SERIAL # 80-001027 LOCATION LHCPD MAIN JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC. Result 0.101 AC.

Instrument is operating properly and accurately. YES (✓) NO ____

COMMENTS END OF STANDARD

SIGNATURE K.C. DRELLER #121
K.C. DRELLER #121
DETECTIVE

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:09:40

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,FREE
DOB: 02/03/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:12:50 |
| Diagnostic Test | Pass | 10:13:24 |
| Air Blank | 0.000 | 10:13:52 |
| 0.100 Cal Check | 0.101 | 10:14:12 |
| Air Blank | 0.000 | 10:14:42 |
| Subject Test | 0.000 | 10:17:43 |
| Air Blank | 0.000 | 10:18:11 |
| Wait | ABT* | |
| Air Blank | 0.000 | 10:22:12 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

AD 121

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:09:40

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,FREE
DOB: 02/03/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:12:50 |
| Diagnostic Test | Pass | 10:13:24 |
| Air Blank | 0.000 | 10:13:52 |
| 0.100 Cal Check | 0.101 | 10:14:12 |
| Air Blank | 0.000 | 10:14:42 |
| Subject Test | 0.000 | 10:17:43 |
| Air Blank | 0.000 | 10:18:11 |
| Wait | ABT* | |
| Air Blank | 0.000 | 10:22:12 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:09:40

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,FREE
DOB: 02/03/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:12:50 |
| Diagnostic Test | Pass | 10:13:24 |
| Air Blank | 0.000 | 10:13:52 |
| 0.100 Cal Check | 0.101 | 10:14:12 |
| Air Blank | 0.000 | 10:14:42 |
| Subject Test | 0.000 | 10:17:43 |
| Air Blank | 0.000 | 10:18:11 |
| Wait | ABT* | |
| Air Blank | 0.000 | 10:22:12 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

121

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:24:03

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,BAWT
DOB: 02/03/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | IPS* | 10:26:29 |
| Air Blank | 0.000 | 10:26:58 |

*Improper Sample

Not a Successfully Completed Test Sequence

121

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:24:03

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,BAWT
DOB: 02/03/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | IPS* | 10:26:29 |
| Air Blank | 0.000 | 10:26:58 |

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:24:03

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,BAWT
DOB: 02/03/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | IPS* | 10:26:29 |
| Air Blank | 0.000 | 10:26:58 |

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:29:58

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,RFI
DOB: 02/03/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | RFI* | 10:31:31 |
| Air Blank | 0.000 | 10:31:59 |

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:29:58

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,RFI
DOB: 02/03/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | RFI* | 10:31:31 |
| Air Blank | 0.000 | 10:31:59 |

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:29:58

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,RFI
DOB: 02/03/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | RFI* | 10:31:31 |
| Air Blank | 0.000 | 10:31:59 |

*RFI Detect

Not a Successfully Completed Test Sequence

#121

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:34:04

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,DEFSAMP
DOB: 02/03/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:41:08 |
| Diagnostic Test | Pass | 10:41:42 |
| Air Blank | 0.000 | 10:42:10 |
| 0.100 Cal Check | 0.101 | 10:42:31 |
| Air Blank | 0.000 | 10:43:00 |
| Subject Test | 0.036* | 10:46:07 |
| Air Blank | 0.000 | 10:46:40 |
| Air Blank | 0.000 | 10:47:07 |
| Subject Test | XXX** | 10:47:54 |
| Air Blank | 0.000 | 10:48:28 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:34:04

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,DEFSAMP
DOB: 02/03/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:41:08 |
| Diagnostic Test | Pass | 10:41:42 |
| Air Blank | 0.000 | 10:42:10 |
| 0.100 Cal Check | 0.101 | 10:42:31 |
| Air Blank | 0.000 | 10:43:00 |
| Subject Test | 0.036* | 10:46:07 |
| Air Blank | 0.000 | 10:46:40 |
| Air Blank | 0.000 | 10:47:07 |
| Subject Test | XXX** | 10:47:54 |
| Air Blank | 0.000 | 10:48:28 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:34:04

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

Subject: SQAP,DEFSAMP
DOB: 02/03/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:41:08 |
| Diagnostic Test | Pass | 10:41:42 |
| Air Blank | 0.000 | 10:42:10 |
| 0.100 Cal Check | 0.101 | 10:42:31 |
| Air Blank | 0.000 | 10:43:00 |
| Subject Test | 0.036* | 10:46:07 |
| Air Blank | 0.000 | 10:46:40 |
| Air Blank | 0.000 | 10:47:07 |
| Subject Test | XXX** | 10:47:54 |
| Air Blank | 0.000 | 10:48:28 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K.C. Deller #121 AGENCY Lake Havasu City Police Dept

DATE 02 03 2011 TIME 1052

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD Main Jail

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 1000 AC.

   OR

   Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.101 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K.C. Deller #121

K.C. Deller #121
Detective

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
02/03/2011 10:52:24

---

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

---

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:53:05 |
| 0.100 Cal Check | 0.101 | 10:53:26 |
| Air Blank | 0.000 | 10:53:55 |

Preventative Maintenance
Performed -- Timer Reset