POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Earl Wilson is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 30 day of MARCH.

2012 by CARLA DUARTE

NOTARY PUBLIC

3-30-12

DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2013

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
TDD 855-4114
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Earl Wilson AGENCY Lake Havasu City Police Dept.

DATE 03/23/12 TIME 09:19:01

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD Jail

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
03/23/2012 09:17:16

Standard Lot#: 13711100A1
Last Changed By: J. NOBLE #5190

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2010
Sex: M Weight: 120

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:19:01
Diagnostic Test	Pass	09:19:35
Air Blank	0.000	09:20:03
0.100 Cal Check	0.099	09:20:23
Air Blank	0.000	09:20:53
Subject Test	0.000	09:21:12
Air Blank	ABT*	09:21:23
Air Blank	IPS**	09:21:32

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

EW 226

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
03/23/2012 09:17:16

Standard Lot#: 13711100A1
Last Changed By: J. NOBLE #5190

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2010
Sex: M Weight: 120

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:19:01
Diagnostic Test	Pass	09:19:35
Air Blank	0.000	09:20:03
0.100 Cal Check	0.099	09:20:23
Air Blank	0.000	09:20:53
Subject Test	0.000	09:21:12
Air Blank	ABT*	09:21:23
Air Blank	IPS**	09:21:32

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

EW 226

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
03/23/2012 09:17:16

Standard Lot#: 13711100A1
Last Changed By: J. NOBLE #5190

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2010
Sex: M Weight: 120

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:19:01
Diagnostic Test	Pass	09:19:35
Air Blank	0.000	09:20:03
0.100 Cal Check	0.099	09:20:23
Air Blank	0.000	09:20:53
Subject Test	0.000	09:21:12
Air Blank	ABT*	09:21:23
Air Blank	IPS**	09:21:32

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

EW 226

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
03/23/2012 09:22:55

Standard Lot#: 13711100A1
Last Changed By: J. NOBLE #5190

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2010
Sex: M Weight: 120

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:24:01
Diagnostic Test	Pass	09:24:35
Air Blank	0.000	09:25:02
0.100 Cal Check	0.100	09:25:23
Air Blank	0.000	09:25:53
Subject Test	0.585*	09:29:00
Air Blank	0.000	09:29:34
Air Blank	0.000	09:30:01
Subject Test	XXX**	09:30:24
Air Blank	RFI***	09:30:35

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EW 226

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
03/23/2012 09:22:55

Standard Lot#: 13711100A1
Last Changed By: J. NOBLE #5190

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2010
Sex: M Weight: 120

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:24:01
Diagnostic Test	Pass	09:24:35
Air Blank	0.000	09:25:02
0.100 Cal Check	0.100	09:25:23
Air Blank	0.000	09:25:53
Subject Test	0.585*	09:29:00
Air Blank	0.000	09:29:34
Air Blank	0.000	09:30:01
Subject Test	XXX**	09:30:24
Air Blank	RFI***	09:30:35

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EW 226

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
03/23/2012 09:22:55

Standard Lot#: 13711100A1
Last Changed By: J. NOBLE #5190

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/2010
Sex: M Weight: 120

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:24:01
Diagnostic Test	Pass	09:24:35
Air Blank	0.000	09:25:02
0.100 Cal Check	0.100	09:25:23
Air Blank	0.000	09:25:53
Subject Test	0.585*	09:29:00
Air Blank	0.000	09:29:34
Air Blank	0.000	09:30:01
Subject Test	XXX**	09:30:24
Air Blank	RFI***	09:30:35

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXO 226

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST EARL WILSON AGENCY LAKE HAVASU CITY POLICE DEPT. JAIL

DATE 03/23/12 TIME 09:11:51

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [signature] 226

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
03/23/2012 09:11:12

Standard Lot#: 13711100A1
Last Changed By: J. NOBLE #5190

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	09:11:51
0.100 Cal Check	0.099	09:12:12
Air Blank	0.000	09:12:41

Preventative Maintenance
Performed -- Timer Reset

EW 226

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST EARL WILSON AGENCY LAKE HAVASU CITY POLICE DEPT.

DATE 03/23/12 TIME 09:36:06

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [signature] 226

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
03/23/2012 09:35:30

Standard Lot#: 13711100A1
Last Changed By: J. NOBLE #5190

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	09:36:06
0.100 Cal Check	0.097	09:36:27
Air Blank	0.000	09:36:57

Preventative Maintenance
Performed -- Timer Reset

EW 226