POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

J Sondrol 187
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 9 day of APRIL,
2013 by CARLA DUARTE

Carla Duarte
NOTARY PUBLIC

4-9-13
DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2017

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	*TDD*	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

187

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LNC PD

DATE 4-8-13 TIME 2053

INTOXILYZER SERIAL # 80-001873 LOCATION LB CHANNEL SIGN 8

( ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

( ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

( ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

( ) 4. Air blank completed.

( ) 5. Calibration check completed. Test results 0.098 AC.

( ) 6. Air blank completed.

( ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

#187

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-101873
Core Version: 8105.48
04/08/2013 20:53:01

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

DAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	20:53:35
0.100 Cal Check	0.096	20:53:56
Air Blank	0.000	20:54:26

Preventative Maintenance
Performed -- Timer Reset

4187

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J Sondrol 187 AGENCY LNC PD

DATE 4-8-13 TIME 2057

INTOXILYZER SERIAL # 80-001873 LOCATION LB CHANNEL SIGN 8

(Y) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(Y) 1. Clock time check.
(Y) 2. Date check.

OPERATIONAL TESTS
(Y) 1. Alcohol-free subject test result 0. 000 AC.
(Y) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(Y) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(Y) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE J Sondrol 187

DPS Form Exh G-4 (Rev 05-01)

#187

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001873
Core Version: 8105.48
04/08/2013 21:01:16

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

DAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SOAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	21:02:05
Diagnostic Test	Pass	21:02:42
Air Blank	0.000	21:03:10
0.100 Cal Check	0.097	21:03:30
Air Blank	0.000	21:04:01
Subject Test	0.316*	21:07:09
Air Blank	0.000	21:17:40
Air Blank	0.000	21:08:08
Subject Test	XXX**	21:08:26
Air Blank	RFI***	21:06:34

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed

#187

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001873
Core Version: 8105.48
04/08/2013 20:57:03

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

DAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SOAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	20:57:51
Diagnostic Test	Pass	20:58:25
Air Blank	0.000	20:58:5
0.100 Cal Check	0.098	20:59:1
Air Blank	0.000	20:59:4
Subject Test	0.000	21:00:0
Air Blank	ABT*	21:00:1
Air Blank	IPS**	21:00:1

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SANDROCK 187 AGENCY LNC PD

DATE 4-8-13 TIME 2110

INTOXILYZER SERIAL # 80-001873 LOCATION CB CHANNEL SIGN 8

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sandrock 187

DPS Form Exh G-2 (Rev 05-01)

9/187

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001873
Core Version: 8105.48
04/08/2013 21:10:28

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

DAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	21:10:58
0.100 Cal Check	0.097	21:11:19
Air Blank	0.000	21:11:49

Preventative Maintenance
Performed -- Timer Reset