POLICE DEPARTMENT
**LAKE HAVASU CITY**
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA )
) ss
COUNTY OF MOHAVE )

Subscribed and sworn to before me this 17 day of April, 2014 by Carla Duarte

Carla Duarte
NOTARY PUBLIC

4-17-14
DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2017

Area Code 928

| | | | | | |
|---|---|---|---|---|---|
| Administration | 855-4884 | Fax #'s | | Emergency | 9-1-1 |
| Business Office | 855-1171 | Administration | 680-5430 | Non-Emergency Dispatch/ | 855-4111 |
| Investigations | 855-5775 | Business Office | 680-5431 | *TDD* | 855-4114 |
| Patrol | 855-0515 | Investigations/Patrol | 680-5432 | Havasu Silent Witness | 854-TIPS |

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 04-13-14 TIME 1112 HRS

INTOXILYZER SERIAL # 80-003356 LOCATION CHANNEL INTOX SIGN 8

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.098 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [signature] *219

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
04/13/2014 11:11:51

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

---

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:12:31 |
| 0.100 Cal Check | 0.098 | 11:12:52 |
| Air Blank | 0.000 | 11:13:22 |

Preventative Maintenance
Performed -- Timer Reset

**EXHIBIT G-4**
**THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)**

**ARIZONA DEPARTMENT OF PUBLIC SAFETY**

**STANDARD QUALITY ASSURANCE PROCEDURES**
**INTOXILYZER MODEL 8000**

**STANDARD QUALITY ASSURANCE PROCEDURE**

QA SPECIALIST CHARLES TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 04-13-14 TIME 1127

INTOXILYZER SERIAL # 80-003356 LOCATION BRIDGEWATER CHANNEL # 8 INTOX

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE [signature]

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
04/13/2014 11:15:58

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:17:00 |
| Diagnostic Test | Pass | 11:17:35 |
| Air Blank | 0.000 | 11:18:02 |
| 0.100 Cal Check | 0.097 | 11:18:23 |
| Air Blank | 0.000 | 11:18:52 |
| Subject Test | 0.000 | 11:19:15 |
| Air Blank | ABT* | 11:19:24 |
| Air Blank | IPS** | 11:19:32 |

*Sequence Aborted
**Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
04/13/2014 11:15:58

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

---

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:17:00 |
| Diagnostic Test | Pass | 11:17:35 |
| Air Blank | 0.000 | 11:18:02 |
| 0.100 Cal Check | 0.097 | 11:18:23 |
| Air Blank | 0.000 | 11:18:52 |
| Subject Test | 0.000 | 11:19:15 |
| Air Blank | ABT* | 11:19:24 |
| Air Blank | IPS** | 11:19:32 |

*Sequence Aborted
**Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
04/13/2014 11:15:58

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

---

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:17:00 |
| Diagnostic Test | Pass | 11:17:35 |
| Air Blank | 0.000 | 11:18:02 |
| 0.100 Cal Check | 0.097 | 11:18:23 |
| Air Blank | 0.000 | 11:18:52 |
| Subject Test | 0.000 | 11:19:15 |
| Air Blank | ABT* | 11:19:24 |
| Air Blank | IPS** | 11:19:32 |

*Sequence Aborted
**Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
04/13/2014 11:20:11

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:21:05 |
| Diagnostic Test | Pass | 11:21:40 |
| Air Blank | 0.000 | 11:22:08 |
| 0.100 Cal Check | 0.098 | 11:22:28 |
| Air Blank | 0.000 | 11:22:58 |
| Subject Test | 0.048* | 11:26:05 |
| Air Blank | 0.000 | 11:26:36 |
| Air Blank | 0.000 | 11:27:04 |
| Subject Test | XXX** | 11:27:32 |
| Air Blank | RFI*** | 11:27:47 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
04/13/2014 11:20:11

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:21:05 |
| Diagnostic Test | Pass | 11:21:40 |
| Air Blank | 0.000 | 11:22:08 |
| 0.100 Cal Check | 0.098 | 11:22:28 |
| Air Blank | 0.000 | 11:22:58 |
| Subject Test | 0.048* | 11:26:05 |
| Air Blank | 0.000 | 11:26:36 |
| Air Blank | 0.000 | 11:27:04 |
| Subject Test | XXX** | 11:27:32 |
| Air Blank | RFI*** | 11:27:47 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
04/13/2014 11:20:11

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:21:05 |
| Diagnostic Test | Pass | 11:21:40 |
| Air Blank | 0.000 | 11:22:08 |
| 0.100 Cal Check | 0.098 | 11:22:28 |
| Air Blank | 0.000 | 11:22:58 |
| Subject Test | 0.048* | 11:26:05 |
| Air Blank | 0.000 | 11:26:36 |
| Air Blank | 0.000 | 11:27:04 |
| Subject Test | XXX** | 11:27:32 |
| Air Blank | RFI*** | 11:27:47 |

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 04-13-14 TIME

INTOXILYZER SERIAL # 80-003356 LOCATION BRIDGEWATER CHANNEL #8 INTOX

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [signature] #29

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
04/13/2014 11:32:17

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

---

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 11:32:47 |
| 0.100 Cal Check | 0.097 | 11:33:08 |
| Air Blank | 0.000 | 11:33:38 |

Preventative Maintenance
Performed -- Timer Reset