



POLICE DEPARTMENT

LAKE HAVASU CITY

2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Josh Sondrol
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)


Subscribed and sworn to before me this 23 day of APRIL.

2012 by CARLA DUARTE

Carla Duarte
NOTARY PUBLIC

4-23-12
DATE

Commission expires:



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
 TDD 855-4114
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. SONDROZ 187 AGENCY LAKE NAVASU CITY PD

DATE 4-23-12 TIME 1541

INTOXILYZER SERIAL # 80-003355 LOCATION STORAGE (CHANNEL SIGN 8)

- (S) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (S) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (S) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (S) 4. Air blank completed.
- (S) 5. Calibration check completed. Test results 0. 098 AC.
- (S) 6. Air blank completed.
- (S) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Sondroz 187

DPS Form Exh G-2 (Rev 05-01)

* AFTER G-2, I CHANGED THE "LOCATION" TO LB CHANNEL SIGN 8
FROM STORAGE

GS 18/

INTOXILYZER 8000
Location: STORAGE (CHANNEL) (SIGN 8)
Serial Number: 80-003355
Core Version: 8105.48
04/23/2012

15:41:17

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	15:41:51
0.100 Cal Check	0.098	15:42:12
Air Blank	0.000	15:42:42

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J SANDROL 187 AGENCY LAKE NAHVALU CITY PD

DATE 4-23-12 TIME 1544

INTOXILYZER SERIAL # 80-003355 LOCATION LB CHANNEL SIGN 8

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE J Sandrol 187

9/187

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
04/23/2012

15:44:20

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:45:14
Diagnostic Test	Pass	15:45:49
Air Blank	0.000	15:46:17
0.100 Cal Check	0.098	15:46:37
Air Blank	0.000	15:47:07
Subject Test	0.000	15:47:27
Air Blank	ABT*	15:47:34
Air Blank	IPS**	15:47:41

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

9/187

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
04/23/2012

15:47:47

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:48:37
Diagnostic Test	Pass	15:49:12
Air Blank	0.000	15:49:40
0.100 Cal Check	0.098	15:50:00
Air Blank	0.000	15:50:30
Subject Test	0.403*	15:53:37
Air Blank	0.000	15:54:11
Air Blank	0.000	15:54:38
Subject Test	XXX**	15:54:56
Air Blank	RFI***	15:55:08

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDRER 187 AGENCY LAKE HAVASU CITY PD

DATE 4-23-12 TIME 1557

INTOXILYZER SERIAL # 80-003355 LOCATION LB CHANNEL SIGN 8

- (J) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution . 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (J) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (J) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (J) 4. Air blank completed.
- (J) 5. Calibration check completed. Test results 0. 096 AC.
- (J) 6. Air blank completed.
- (J) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrer 187

8187

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
04/23/2012 15:57:21

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	15:57:51
0.100 Cal Check	0.096	15:58:12
Air Blank	0.000	15:58:42

Preventative Maintenance
Performed -- Timer Reset