POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 28 **day of** April.

2014 **by** Carla Duarte

Carla Duarte

NOTARY PUBLIC

4-28-14

DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2017

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	*TDD*	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LHCPD

DATE 04-25-14 TIME _____

INTOXILYZER SERIAL # 80-003355 LOCATION MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.101 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
04/25/2014 11:20:33

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	11:21:03
0.100 Cal Check	0.101	11:21:24
Air Blank	0.000	11:21:53

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LHCPD

DATE 04-25-14 TIME _____

INTOXILYZER SERIAL # 80-00 3355 LOCATION MAIN JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0._____AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC. Result 0.101 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

_____

_____

_____

_____

SIGNATURE _____

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
04/25/2014 11:06:52

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:07:49
Diagnostic Test	Pass	11:08:24
Air Blank	0.000	11:08:51
0.100 Cal Check	0.101	11:09:12
Air Blank	0.000	11:09:41
Subject Test	0.000	11:09:56
Air Blank	ABT*	11:10:19
Air Blank	IPS**	11:10:27

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
04/25/2014 11:06:52

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:07:49
Diagnostic Test	Pass	11:08:24
Air Blank	0.000	11:08:51
0.100 Cal Check	0.101	11:09:12
Air Blank	0.000	11:09:41
Subject Test	0.000	11:09:56
Air Blank	ABT*	11:10:19
Air Blank	IPS**	11:10:27

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
04/25/2014 11:06:52

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:07:49
Diagnostic Test	Pass	11:08:24
Air Blank	0.000	11:08:51
0.100 Cal Check	0.101	11:09:12
Air Blank	0.000	11:09:41
Subject Test	0.000	11:09:56
Air Blank	ABT*	11:10:19
Air Blank	IPS**	11:10:27

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
04/25/2014 11:11:16

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:12:07
Diagnostic Test	Pass	11:12:42
Air Blank	0.000	11:13:10
0.100 Cal Check	0.101	11:13:30
Air Blank	0.000	11:14:00
Subject Test	0.046*	11:17:07
Air Blank	0.000	11:17:38
Air Blank	0.000	11:18:06
Subject Test	XXX**	11:18:23
Air Blank	RFI***	11:18:31

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
04/25/2014 11:11:16

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:12:07
Diagnostic Test	Pass	11:12:42
Air Blank	0.000	11:13:10
0.100 Cal Check	0.101	11:13:30
Air Blank	0.000	11:14:00
Subject Test	0.046*	11:17:07
Air Blank	0.000	11:17:38
Air Blank	0.000	11:18:06
Subject Test	XXX**	11:18:23
Air Blank	RFI***	11:18:31

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
04/25/2014 11:11:16

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:12:07
Diagnostic Test	Pass	11:12:42
Air Blank	0.000	11:13:10
0.100 Cal Check	0.101	11:13:30
Air Blank	0.000	11:14:00
Subject Test	0.046*	11:17:07
Air Blank	0.000	11:17:38
Air Blank	0.000	11:18:06
Subject Test	XXX**	11:18:23
Air Blank	RFI***	11:18:31

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LHCPD

DATE 04-25-14 TIME 1049

INTOXILYZER SERIAL # 80-003355 LOCATION MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
04/25/2014 10:49:58

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	10:51:24
0.100 Cal Check	0.100	10:51:45
Air Blank	0.000	10:52:15

Preventative Maintenance
Performed -- Timer Reset