POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Earl Wilson is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

E. L. Wilson 226

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 8 day of MAY, 2013 by CARLA DUARTE

Carla Duarte

NOTARY PUBLIC

5-8-13

DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2017

Area Code 928

Administration 855-4884	Fax #'s	Emergency 9-1-1
Business Office 855-1171	Administration 680-5430	Non-Emergency Dispatch/ 855-4111
Investigations 855-5775	Business Office 680-5431	*TDD* 855-4114
Patrol 855-0515	Investigations/Patrol 680-5432	Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Earl Wilson AGENCY Lake Havasu City Police Dept.

DATE 05/08/13 TIME 09:14:52

INTOXILYZER SERIAL # 80-003356 LOCATION LHCPD JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE EQ LP 226

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
05/08/2013 09:14:52

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	09:15:29
0.100 Cal Check	0.099	09:15:49
Air Blank	0.000	09:16:19

Preventative Maintenance
Performed -- Timer Reset

EW 226

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST EARL WILSON AGENCY LAKE HAVASU CITY POLICE DEPT.

DATE 05/08/13 TIME 09:23:22

INTOXILYZER SERIAL # 80-003356 LOCATION LHCPD JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC. Result 0.099 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE Earl Wilson 224

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
05/08/2013 09:23:22

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 12/12/2012
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:24:49
Diagnostic Test	Pass	09:25:24
Air Blank	0.000	09:25:51
0.100 Cal Check	0.099	09:26:12
Air Blank	0.000	09:26:42
Subject Test	0.000	09:27:01
Air Blank	ABT*	09:27:08
Air Blank	IPS**	09:27:17

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
05/08/2013 09:29:29

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 12/12/2012
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:30:36
Diagnostic Test	Pass	09:31:11
Air Blank	0.000	09:31:38
0.100 Cal Check	0.099	09:31:59
Air Blank	0.000	09:32:28
Subject Test	0.095*	09:35:36
Air Blank	0.000	09:36:08
Air Blank	0.000	09:36:35
Subject Test	XXX**	09:37:00
Air Blank	RFI***	09:37:08

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST EARL WILSON AGENCY LAKE HAVASU CITY POLICE DEPT.

DATE 05/06/13 TIME 09:40:25

INTOXILYZER SERIAL # 80-003356 LOCATION LHCPD JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(✓) 4. Air blank completed.
(✓) 5. Calibration check completed. Test results 0.099 AC.
(✓) 6. Air blank completed.
(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE EL Wilson 226

DPS Form Exh G-2 (Rev 05-01)

'OXILYZER 8000
:ation: LHC MAIN JAIL
:ial Number: 80-003356
:e Version: 8105.48
'08/2013 09:40:25

---

ndard Lot#: 01513100A1
st Changed By: J. SONDROL #187

3: R. WILSON #226
(E HAVASU CITY PD

---

rator: E. WILSON #226
(E HAVASU CITY PD

---

st	g/210L	Time
: Blank	0.000	09:40:56
100 Cal Check	0.099	09:41:17
: Blank	0.000	09:41:47

Preventative Maintenance
Performed -- Timer Reset

EXW 226