POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

J Sondrol 187
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 21 day of May,
2014 by Carla Duarte

Carla Duarte
NOTARY PUBLIC

5/21/14
DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2017

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	*TDD*	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SANDROL 187 AGENCY LHC PD

DATE 5/20/14 TIME 0132

INTOXILYZER SERIAL # 80-003355 LOCATION MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(✓) 4. Air blank completed.
(✓) 5. Calibration check completed. Test results 0.101 AC.
(✓) 6. Air blank completed.
(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sandrol 187

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
05/20/2014 01:32:35

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	01:33:08
0.100 Cal Check	0.101	01:33:29
Air Blank	0.000	01:33:58

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J SONDRAL 187 AGENCY CNC PD

DATE 5/20/14 TIME 0135

INTOXILYZER SERIAL # 80-003355 LOCATION MAIN JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 101 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE J Sondral 187

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
05/20/2014 01:35:16

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:36:12
Diagnostic Test	Pass	01:36:47
Air Blank	0.000	01:37:14
0.100 Cal Check	0.100	01:37:35
Air Blank	0.000	01:38:05
Subject Test	0.000	01:38:26
Air Blank	ABT*	01:38:33
Air Blank	IPS**	01:38:41

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
05/20/2014 01:35:16

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:36:12
Diagnostic Test	Pass	01:36:47
Air Blank	0.000	01:37:14
0.100 Cal Check	0.100	01:37:35
Air Blank	0.000	01:38:05
Subject Test	0.000	01:38:26
Air Blank	ABT*	01:38:33
Air Blank	IPS**	01:38:41

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
05/20/2014 01:35:16

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:36:12
Diagnostic Test	Pass	01:36:47
Air Blank	0.000	01:37:14
0.100 Cal Check	0.100	01:37:35
Air Blank	0.000	01:38:05
Subject Test	0.000	01:38:26
Air Blank	ABT*	01:38:33
Air Blank	IPS**	01:38:41

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
05/20/2014 01:40:02

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:40:53
Diagnostic Test	Pass	01:41:28
Air Blank	0.000	01:41:55
0.100 Cal Check	0.101	01:42:16
Air Blank	0.000	01:42:45
Subject Test	0.087*	01:45:53
Air Blank	0.000	01:46:28
Air Blank	0.000	01:46:55
Subject Test	XXX**	01:47:21
Air Blank	RFI***	01:47:28

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
05/20/2014 01:40:02

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:40:53
Diagnostic Test	Pass	01:41:28
Air Blank	0.000	01:41:55
0.100 Cal Check	0.101	01:42:16
Air Blank	0.000	01:42:45
Subject Test	0.087*	01:45:53
Air Blank	0.000	01:46:28
Air Blank	0.000	01:46:55
Subject Test	XXX**	01:47:21
Air Blank	RFI***	01:47:28

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
05/20/2014 01:40:02

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:40:53
Diagnostic Test	Pass	01:41:28
Air Blank	0.000	01:41:55
0.100 Cal Check	0.101	01:42:16
Air Blank	0.000	01:42:45
Subject Test	0.087*	01:45:53
Air Blank	0.000	01:46:28
Air Blank	0.000	01:46:55
Subject Test	XXX**	01:47:21
Air Blank	RFI***	01:47:28

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LHC PD

DATE 5/20/14 TIME 0149

INTOXILYZER SERIAL # 80-003355 LOCATION MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

* UNSTABLE SIGNAL, THEN PM PERFORMED FINE

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
05/20/2014 01:49:13

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	01:49:43
Reference	UNS*	01:49:51
Air Blank	0.000	01:50:19

*Unstable Signal

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
05/20/2014 01:50:37

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	01:51:08
0.100 Cal Check	0.100	01:51:29
Air Blank	0.000	01:51:59

Preventative Maintenance
Performed -- Timer Reset