POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Earl Wilson is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

E L S 226

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 15 day of JUNE, 2012 by CARLA DUARTE

Carla Duarte

NOTARY PUBLIC

6-15-12

DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2013

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	*TDD*	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST EARL WILSON AGENCY LAKE HAVASU CITY POLICE DEPT.

DATE 05/23/12 TIME 10:35:27

INTOXILYZER SERIAL # 80-003355 LOCATION BRIDGEWATER CHANNEL (SIGNAL)

(Y) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(Y) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(Y) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(Y) 4. Air blank completed.

(Y) 5. Calibration check completed. Test results 0.097 AC.

(Y) 6. Air blank completed.

(Y) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE E L C 224

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
05/23/2012 10:34:56

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	10:35:27
0.100 Cal Check	0.097	10:35:48
Air Blank	0.000	10:36:18

Preventative Maintenance
Performed -- Timer Reset

226

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Earl Wilson AGENCY Lake Havasu City Police Dept.

DATE 05/23/12 TIME 10:11:05

INTOXILYZER SERIAL # 80-003355 LOCATION Bridgewater Channel (SIGMA8)

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE Earl Wilson 226

DPS Form Exh G-4 (Rev 05-01)

At 10:22:59 No Alcohol sample completed.

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
05/23/2012 10:26:03

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/2010
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:27:17
Diagnostic Test	Pass	10:27:52
Air Blank	0.000	10:28:19
0.100 Cal Check	0.097	10:28:40
Air Blank	0.000	10:29:10
Subject Test	0.177*	10:32:17
Air Blank	0.000	10:32:50
Air Blank	0.000	10:33:17
Subject Test	XXX**	10:33:39
Air Blank	RFI***	10:33:47

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

BW 226

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
05/23/2012 10:15:31

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/2010
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:16:34
Diagnostic Test	Pass	10:17:09
Air Blank	0.000	10:17:37
0.100 Cal Check	0.098	10:17:57
Air Blank	0.000	10:18:27
Subject Test	0.433*	10:21:34
Air Blank	0.000	10:22:08
Air Blank	0.000	10:22:35
Subject Test	0.036	10:22:59
Air Blank	RFI**	10:23:07
Air Blank	0.000	10:23:35

*Deficient Sample
**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
05/23/2012 10:10:04

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/2010
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:11:05
Diagnostic Test	Pass	10:11:40
Air Blank	0.000	10:12:08
0.100 Cal Check	0.097	10:12:28
Air Blank	0.000	10:12:58
Subject Test	0.000	10:13:13
Air Blank	ABT*	10:13:22
Air Blank	IPS**	10:13:30

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST EARL WILSON AGENCY LAKE HAVASU CITY POLICE DEPT.

DATE 05/23/12 TIME 10:03:40

INTOXILYZER SERIAL # 80-003355 LOCATION BRIDGEWATER CHANNEL (SIGN #8)

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.098 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [signature] 226

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
05/23/2012 10:03:07

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	10:03:40
0.100 Cal Check	0.098	10:04:01
Air Blank	0.000	10:04:30

Preventative Maintenance
Performed -- Timer Reset

AW 226