POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Karl C. E. Dreller is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are a full, true and correct copy of the original.

K.C.D
SQAP SIGNATURE 05282010
SN#80-001029
CHANGE OF STANDARD

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 4TH day of June, 2010 by K.C. DRELLER.

Gloria Thomas
NOTARY PUBLIC

6-4-10
DATE

Commission expires: 8-31-10

Notary Public State of Arizona
Mohave County
Gloria Thomas
My Commission Expires
08/31/2010

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
TDD 855-4114
Havasu Silent Witness 854-TIPS

E-mail: police@ci.lake-havasu-city.az.us

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K.C. DRELLER #121 AGENCY LAKE HAVASU CITY POLICE DEPT.

DATE 05282010 TIME 1117

INTOXILYZER SERIAL # 80·001029 LOCATION LHCPD MAIN JAIL

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.099 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K.C. Dreller 121

K.C. DRELLER #121
DETECTIVE

DPS Form Exh G-2 (Rev 05-01)

[illegible]

[illegible]

[illegible]	[illegible]	[illegible]
[illegible]	[illegible]	[illegible]
[illegible]	[illegible]	[illegible]
[illegible]	[illegible]	[illegible]

[illegible]

Performed -- [illegible]

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K.C. Deeller #121 AGENCY Lake Havasu City Police Dept

DATE 05 28 2010 TIME 1126

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD Main

(KCD) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(KCD) 1. Clock time check.
(KCD) 2. Date check.

OPERATIONAL TESTS
(KCD) 1. Alcohol-free subject test result 0. 000 AC.
(KCD) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(KCD) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(KCD) 4. Standard Calibration Check standard 0. .100 AC. Result 0. 101 AC.

Instrument is operating properly and accurately. YES (KCD) NO ____

COMMENTS Begin Service Lot # 712322 Exp 010512

SIGNATURE K.C. Deeller #121
K.C. Deeller #121
Detective

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
05/28/2010 11:26:48

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,FREE
DOB: 05/28/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:28:12
Diagnostic Test	Pass	11:28:46
Air Blank	0.000	11:29:13
0.100 Cal Check	0.100	11:29:34
Air Blank	0.000	11:30:04
Subject Test	0.000	11:30:30
Air Blank	0.000	11:30:57
Wait	ABT*	
Air Blank	0.000	11:32:07

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
05/28/2010 11:26:48

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,FREE
DOB: 05/28/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:28:12
Diagnostic Test	Pass	11:28:46
Air Blank	0.000	11:29:13
0.100 Cal Check	0.100	11:29:34
Air Blank	0.000	11:30:04
Subject Test	0.000	11:30:30
Air Blank	0.000	11:30:57
Wait	ABT*	
Air Blank	0.000	11:32:07

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
05/28/2010 11:26:48

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,FREE
DOB: 05/28/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:28:12
Diagnostic Test	Pass	11:28:46
Air Blank	0.000	11:29:13
0.100 Cal Check	0.100	11:29:34
Air Blank	0.000	11:30:04
Subject Test	0.000	11:30:30
Air Blank	0.000	11:30:57
Wait	ABT*	
Air Blank	0.000	11:32:07

*Sequence Aborted

Not a Successfully Completed Test Sequence

KCB

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
05/28/2010 11:34:28

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,BAWT
DOB: 05/28/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	11:37:27
Air Blank	0.000	11:37:56

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
05/28/2010 11:34:28

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,BAWT
DOB: 05/28/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	11:37:27
Air Blank	0.000	11:37:56

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
05/28/2010 11:34:28

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SOAP,BAWT
DOB: 05/28/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	11:37:27
Air Blank	0.000	11:37:56

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
05/28/2010 11:38:49

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,RFI
DOB: 05/28/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	11:39:40
Air Blank	0.000	11:40:09

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
05/28/2010 11:38:49

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,RFI
DOB: 05/28/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	11:39:40
Air Blank	0.000	11:40:09

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
05/28/2010 11:38:49

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,RFI
DOB: 05/28/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	11:39:40
Air Blank	0.000	11:40:09

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
05/28/2010 11:40:39

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,DEFSAMP
DOB: 05/28/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:42:12
Diagnostic Test	Pass	11:42:46
Air Blank	0.000	11:43:13
0.100 Cal Check	0.100	11:43:34
Air Blank	0.000	11:44:03
Subject Test	0.067*	11:47:11
Air Blank	0.000	11:47:43
Air Blank	0.000	11:48:10
Subject Test	XXX**	11:49:10
Air Blank	0.000	11:49:50

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
05/28/2010 11:40:33

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,DEFSAMP
DOB: 05/28/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:42:12
Diagnostic Test	Pass	11:42:46
Air Blank	0.000	11:43:13
0.100 Cal Check	0.100	11:43:34
Air Blank	0.000	11:44:03
Subject Test	0.067*	11:47:11
Air Blank	0.000	11:47:43
Air Blank	0.000	11:48:10
Subject Test	XXX**	11:49:10
Air Blank	0.000	11:49:50

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
05/28/2010 11:40:38

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SQAP,DEFSAMP
DOB: 05/28/2000
Sex: M Weight: 121

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:42:12
Diagnostic Test	Pass	11:42:46
Air Blank	0.000	11:43:13
0.100 Cal Check	0.100	11:43:34
Air Blank	0.000	11:44:03
Subject Test	0.067*	11:47:11
Air Blank	0.000	11:47:43
Air Blank	0.000	11:48:10
Subject Test	XXX**	11:49:10
Air Blank	0.000	11:49:50

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K.C. DEELER #121 AGENCY Lake Havasu City Police Dept

DATE 05282010 TIME 1150

INTOXILYZER SERIAL # 80-001029 LOCATION LCPD Main Jail

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

   OR

   Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.101 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K.C. Deeler 121

K.C. DEELER #121
DETECTIVE

DPS Form Exh G-2 (Rev 05-01)

[illegible]
Location: [illegible]
Serial Number: [illegible]
Core Version: [illegible]
05 28 2010 [illegible]

Standard Bottle: [illegible]
Last Changed [illegible]

QAS: [illegible]
LAKE HAVASU CITY [illegible]

Operator: [illegible]
LAKE HAVASU CITY [illegible]

Test	[illegible]	[illegible]
Air Blank	0.009	[illegible]
0.100 Cal Check	0.101	[illegible]
Air Blank	0.009	[illegible]
Air Blank	0.009	[illegible]

*Input Timed Out