



POLICE DEPARTMENT  
**LAKE HAVASU CITY**  
 2360 McCULLOCH BOULEVARD NORTH  
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Josh Sondrol  
 Signature

STATE OF ARIZONA        )  
   ) ss  
 COUNTY OF MOHAVE     )


Subscribed and sworn to before me this 9 day of June

2015 by Josh Sondrol

Carla Duarte  
 NOTARY PUBLIC

6/9/15  
 DATE

Commission expires:



Area Code 928

Administration ..... 855-4884  
 Business Office ..... 855-1171  
 Investigations ..... 855-5775  
 Patrol ..... 855-0515

Fax #'s  
 Administration ..... 680-5430  
 Business Office ..... 680-5431  
 Investigations/Patrol ..... 680-5432

Emergency ..... 9-1-1  
 Non-Emergency Dispatch/ ..... 855-4111  
   TDD 855-4114  
 Havasu Silent Witness ..... 854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROCK 187 AGENCY LHC PD

DATE 6-4-15 TIME 0527

INTOXILYZER SERIAL # 80-003356 LOCATION LHC MAIN JAIL

() 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

() 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

() 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

() 4. Air blank completed.

() 5. Calibration check completed. Test results 0.096 AC.

() 6. Air blank completed.

() 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrock 187

INTOXILYZER 8000  
Location: LHC MAIN JAIL  
Serial Number: 80-003356  
Core Version: 8105.48  
06/04/2015

05:27:36

-----  
Standard Lot#: 01314100A3  
Last Changed By: J. SONDROL #187  
-----

QAS: J. SONDROL #187  
LAKE HAVASU CITY PD  
-----

Operator: J. SONDROL #187  
LAKE HAVASU CITY PD  
-----

Test	g/210L	Time
Air Blank	0.000	05:28:09
0.100 Cal Check	0.096	05:28:29
Air Blank	0.000	05:28:59

Preventative Maintenance  
Performed -- Timer Reset