POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 2 day of July, 2014 by Charles Trescher

R Shoemake
NOTARY PUBLIC

7/2/14
DATE

Commission expires: July 13, 2015

ROBIN SHOEMAKE
Notary Public - Arizona
Mohave County
My Comm. Expires Jul 13, 2015

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
TDD 855-4114
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

#219

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LHCPD

DATE 6-28-14 TIME 1324 HRS

INTOXILYZER SERIAL # 80-003356 LOCATION CHANNEL INTOX

( ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

( ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

( ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

( ) 4. Air blank completed.

( ) 5. Calibration check completed. Test results 0.096 AC.

( ) 6. Air blank completed.

( ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

#219

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
06/28/2014 13:24:17

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	13:24:52
0.100 Cal Check	0.096	13:25:13
Air Blank	0.000	13:25:42

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LHCPD

DATE 6-28-14 TIME 1331

INTOXILYZER SERIAL # 80-003356 LOCATION CHANNEL INTOX

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC. Result 0.097 AC.

Instrument is operating properly and accurately. YES X NO ____

COMMENTS

SIGNATURE

DPS Form Exh G-4 (Rev 05-01)

CT 219

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
06/28/2014 13:31:43

Standard Lot#: 01513100AI
Last Changed By: J. SONDROL #187

OAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SOAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Depriuation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:32:38
Diagnostic Test	Pass	13:33:15
Air Blank	0.000	13:33:42
0.100 Cal Check	0.097	13:34:03
Air Blank	0.000	13:34:33
Subject Test	0.035	13:34:51
Air Blank	0.000	13:35:23
Wait		
Air Blank	0.000	13:40:19
Subject Test	XXX*	13:40:48
Air Blank	RFI**	13:40:55

*Invalid Sample
Begin new depriuation period
**RFI Detect

Not a Successfully Completed Test Sequence

*CT 219

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
06/28/2014 13:31:43

Standard Lot#: 01513100AI
Last Changed By: J. SONDROL #187

OAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SOAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Depriuation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:32:38
Diagnostic Test	Pass	13:33:15
Air Blank	0.000	13:33:42
0.100 Cal Check	0.097	13:34:03
Air Blank	0.000	13:34:33
Subject Test	0.035	13:34:51
Air Blank	0.000	13:35:23
Wait		
Air Blank	0.000	13:40:19
Subject Test	XXX*	13:40:48
Air Blank	RFI**	13:40:55

*Invalid Sample
Begin new depriuation period
**RFI Detect

Not a Successfully Completed Test Sequence

CT 219

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
06/28/2014 13:31:43

Standard Lot#: 01513100AI
Last Changed By: J. SONDROL #187

OAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SOAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Depriuation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:32:38
Diagnostic Test	Pass	13:33:15
Air Blank	0.000	13:33:42
0.100 Cal Check	0.097	13:34:03
Air Blank	0.000	13:34:33
Subject Test	0.035	13:34:51
Air Blank	0.000	13:35:23
Wait		
Air Blank	0.000	13:40:19
Subject Test	XXX*	13:40:48
Air Blank	RFI**	13:40:55

*Invalid Sample
Begin new depriuation period
**RFI Detect

Not a Successfully Completed Test Sequence

#219

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
06/28/2014 13:27:20

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:28:21
Diagnostic Test	Pass	13:28:56
Air Blank	0.000	13:29:24
0.100 Cal Check	0.097	13:29:44
Air Blank	0.000	13:30:14
Subject Test	0.000	13:30:33
Air Blank	ABT*	13:30:39
Air Blank	IPS**	13:30:47

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

C1
#219

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
06/28/2014 13:27:20

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:28:21
Diagnostic Test	Pass	13:28:56
Air Blank	0.000	13:29:24
0.100 Cal Check	0.097	13:29:44
Air Blank	0.000	13:30:14
Subject Test	0.000	13:30:33
Air Blank	ABT*	13:30:39
Air Blank	IPS**	13:30:47

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

#219

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
06/28/2014 13:27:20

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:28:21
Diagnostic Test	Pass	13:28:56
Air Blank	0.000	13:29:24
0.100 Cal Check	0.097	13:29:44
Air Blank	0.000	13:30:14
Subject Test	0.000	13:30:33
Air Blank	ABT*	13:30:39
Air Blank	IPS**	13:30:47

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LNCPD

DATE 6-28-14 TIME 1344

INTOXILYZER SERIAL # 80-003356 LOCATION CHANNEL INTOX

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.096 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003356
Core Version: 8105.48
06/28/2014 13:44:47

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	13:45:18
0.100 Cal Check	0.096	13:45:39
Air Blank	0.000	13:46:09

Preventative Maintenance
Performed -- Timer Reset

POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 2 day of July.
2014 by Charles Trescher
Rshoemake
NOTARY PUBLIC

7/2/14
DATE

Commission expires:
July 13, 2015

ROBIN SHOEMAKE
Notary Public - Arizona
Mohave County
My Comm. Expires Jul 13, 2015

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	*TDD*	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LHCPD

DATE 6-28-14 TIME 1039 HRS

INTOXILYZER SERIAL # 80-003355 LOCATION LHCPD MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

#219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
06/28/2014 10:39:39

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	10:40:21
0.100 Cal Check	0.100	10:40:42
Air Blank	0.000	10:41:12

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LMCPD

DATE 6-28-14 TIME 1048

INTOXILYZER SERIAL # 80-003355 LOCATION LMCPD MAIN JAIL

(√) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(√) 1. Clock time check.
(√) 2. Date check.

OPERATIONAL TESTS
(√) 1. Alcohol-free subject test result 0. 000 AC.
(√) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(√) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(√) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 101 AC.

Instrument is operating properly and accurately. YES √ NO ____

COMMENTS

SIGNATURE [signature]

DPS Form Exh G-4 (Rev 05-01)

*219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
06/28/2014 10:43:16

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:44:24
Diagnostic Test	Pass	10:44:59
Air Blank	0.000	10:45:26
0.100 Cal Check	0.101	10:45:47
Air Blank	0.000	10:46:17
Subject Test	0.000	10:46:34
Air Blank	ABT*	10:46:41
Air Blank	IPS**	10:46:49

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

#219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
06/28/2014 10:43:16

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:44:24
Diagnostic Test	Pass	10:44:59
Air Blank	0.000	10:45:26
0.100 Cal Check	0.101	10:45:47
Air Blank	0.000	10:46:17
Subject Test	0.000	10:46:34
Air Blank	ABT*	10:46:41
Air Blank	IPS**	10:46:49

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

#219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
06/28/2014 10:43:16

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:44:24
Diagnostic Test	Pass	10:44:59
Air Blank	0.000	10:45:26
0.100 Cal Check	0.101	10:45:47
Air Blank	0.000	10:46:17
Subject Test	0.000	10:46:34
Air Blank	ABT*	10:46:41
Air Blank	IPS**	10:46:49

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
06/28/2014 10:48:00

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:48:53
Diagnostic Test	Pass	10:49:28
Air Blank	0.000	10:49:56
0.100 Cal Check	0.101	10:50:16
Air Blank	0.000	10:50:46
Subject Test	0.036*	10:53:53
Air Blank	0.000	10:54:24
Air Blank	0.000	10:54:52
Subject Test	XXX**	10:55:12
Air Blank	RFI***	10:55:22

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

#219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
06/28/2014 10:48:00

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:48:53
Diagnostic Test	Pass	10:49:28
Air Blank	0.000	10:49:56
0.100 Cal Check	0.101	10:50:16
Air Blank	0.000	10:50:46
Subject Test	0.036*	10:53:53
Air Blank	0.000	10:54:24
Air Blank	0.000	10:54:52
Subject Test	XXX**	10:55:12
Air Blank	RFI***	10:55:22

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
06/28/2014 10:48:00

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:48:53
Diagnostic Test	Pass	10:49:28
Air Blank	0.000	10:49:56
0.100 Cal Check	0.101	10:50:16
Air Blank	0.000	10:50:46
Subject Test	0.036*	10:53:53
Air Blank	0.000	10:54:24
Air Blank	0.000	10:54:52
Subject Test	XXX**	10:55:12
Air Blank	RFI***	10:55:22

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LHCPD

DATE 6-28-14 TIME 1057 HRS

INTOXILYZER SERIAL # 80-003355 LOCATION LHCPD MAIN

(√) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(√) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(√) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(√) 4. Air blank completed.

(√) 5. Calibration check completed. Test results 0. 102 AC.

(√) 6. Air blank completed.

(√) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003355
Core Version: 8105.48
06/28/2014 10:57:08

Standard Lot#: 01513100A1
Last Changed By: C. TRESCHER #219

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	10:57:39
0.100 Cal Check	0.102	10:58:00
Air Blank	0.000	10:58:30

Preventative Maintenance
Performed -- Timer Reset