POLICE DEPARTMENT
**LAKE HAVASU CITY**
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA )
) ss
COUNTY OF MOHAVE )

Subscribed and sworn to before me this 29 day of JULY.
2013 by CARLA DUARTE

NOTARY PUBLIC

7/29/13
DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2017

Area Code 928

Administration ..... 855-4884
Business Office ..... 855-1171
Investigations ..... 855-5775
Patrol ..... 855-0515

Fax #'s
Administration ..... 680-5430
Business Office ..... 680-5431
Investigations/Patrol ..... 680-5432

Emergency ..... 9-1-1
Non-Emergency Dispatch/ ..... 855-4111
TDD 855-4114
Havasu Silent Witness ..... 854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LHC PD

DATE 7/26/13 TIME 1556

INTOXILYZER SERIAL # 80-003356 LOCATION LHC MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution , 100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

p'0 /

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
07/26/2013 15:56:59

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 15:57:31 |
| 0.100 Cal Check | 0.097 | 15:57:52 |
| Air Blank | 0.000 | 15:58:21 |

Preventative Maintenance
Performed -- Timer Reset

YS187

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LNC PD

DATE 7/26/13 TIME 1600

INTOXILYZER SERIAL # 80-003356 LOCATION LNC MAIN JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE J Sondrol 187

DPS Form Exh G-4 (Rev 05-01)

g187

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
07/26/2013 16:00:15

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 16:01:32 |
| Diagnostic Test | Pass | 16:02:08 |
| Air Blank | 0.000 | 16:02:35 |
| 0.100 Cal Check | 0.098 | 16:02:56 |
| Air Blank | 0.000 | 16:03:26 |
| Subject Test | 0.000 | 16:03:42 |
| Air Blank | ABT* | 16:03:48 |
| Air Blank | IPS** | 16:03:56 |

*Sequence Aborted
**Improper Sample

**Not a Successfully Completed Test Sequence**

918

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
07/26/2013 16:00:15

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 16:01:32 |
| Diagnostic Test | Pass | 16:02:08 |
| Air Blank | 0.000 | 16:02:35 |
| 0.100 Cal Check | 0.098 | 16:02:56 |
| Air Blank | 0.000 | 16:03:26 |
| Subject Test | 0.000 | 16:03:42 |
| Air Blank | ABT* | 16:03:48 |
| Air Blank | IPS** | 16:03:56 |

*Sequence Aborted
**Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
07/26/2013 16:00:15

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 16:01:32 |
| Diagnostic Test | Pass | 16:02:08 |
| Air Blank | 0.000 | 16:02:35 |
| 0.100 Cal Check | 0.098 | 16:02:56 |
| Air Blank | 0.000 | 16:03:26 |
| Subject Test | 0.000 | 16:03:42 |
| Air Blank | ABT* | 16:03:48 |
| Air Blank | IPS** | 16:03:56 |

*Sequence Aborted
**Improper Sample

**Not a Successfully Completed Test Sequence**

p 181

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
07/26/2013 16:04:23

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 16:05:13 |
| Diagnostic Test | Pass | 16:05:48 |
| Air Blank | 0.000 | 16:06:15 |
| 0.100 Cal Check | 0.097 | 16:06:36 |
| Air Blank | 0.000 | 16:07:06 |
| Subject Test | XXX* | 16:07:23 |
| Air Blank | 0.000 | 16:07:56 |

*Invalid Sample
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
07/26/2013 16:04:23

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 16:05:13 |
| Diagnostic Test | Pass | 16:05:48 |
| Air Blank | 0.000 | 16:06:15 |
| 0.100 Cal Check | 0.097 | 16:06:36 |
| Air Blank | 0.000 | 16:07:06 |
| Subject Test | XXX* | 16:07:23 |
| Air Blank | 0.000 | 16:07:56 |

*Invalid Sample
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
07/26/2013 16:04:23

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 16:05:13 |
| Diagnostic Test | Pass | 16:05:48 |
| Air Blank | 0.000 | 16:06:15 |
| 0.100 Cal Check | 0.097 | 16:06:36 |
| Air Blank | 0.000 | 16:07:06 |
| Subject Test | XXX* | 16:07:23 |
| Air Blank | 0.000 | 16:07:56 |

*Invalid Sample
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

7187

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
07/26/2013 16:08:24

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 16:09:17 |
| Diagnostic Test | Pass | 16:09:51 |
| Air Blank | 0.000 | 16:10:19 |
| 0.100 Cal Check | 0.097 | 16:10:40 |
| Air Blank | 0.000 | 16:11:09 |
| Subject Test | 0.350* | 16:14:17 |
| Air Blank | RFI** | 16:14:25 |
| Air Blank | 0.000 | 16:14:53 |

*Deficient Sample
**RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
07/26/2013 16:08:24

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 16:09:17 |
| Diagnostic Test | Pass | 16:09:51 |
| Air Blank | 0.000 | 16:10:19 |
| 0.100 Cal Check | 0.097 | 16:10:40 |
| Air Blank | 0.000 | 16:11:09 |
| Subject Test | 0.350* | 16:14:17 |
| Air Blank | RFI** | 16:14:25 |
| Air Blank | 0.000 | 16:14:53 |

*Deficient Sample
**RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
07/26/2013 16:08:24

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 16:09:17 |
| Diagnostic Test | Pass | 16:09:51 |
| Air Blank | 0.000 | 16:10:19 |
| 0.100 Cal Check | 0.097 | 16:10:40 |
| Air Blank | 0.000 | 16:11:09 |
| Subject Test | 0.350* | 16:14:17 |
| Air Blank | RFI** | 16:14:25 |
| Air Blank | 0.000 | 16:14:53 |

*Deficient Sample
**RFI Detect

**Not a Successfully Completed Test Sequence**

45187

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LNC PD

DATE 7/26/13 TIME 1615

INTOXILYZER SERIAL # 80-003356 LOCATION LNC MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.096 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
07/26/2013 16:15:33

---

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

---

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 16:16:03 |
| 0.100 Cal Check | 0.096 | 16:16:24 |
| Air Blank | 0.000 | 16:16:54 |

Preventative Maintenance
Performed -- Timer Reset