POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Karl C. E. Dreller is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are a full, true and correct copy of the original.

K. C. Dreller
SQAP SIGNATURE 07292010
SN# 80.001029

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 24 day of SEPTEMBER, 2010 by CARLA DUARTE.

Carla Duarte
NOTARY PUBLIC

9-24-10
DATE

Commission expires: 3/29/2013

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2013

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

E-mail: police@ci.lake-havasu-city.az.us

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
TDD 855-4114
Havasu Silent Witness 854-TIPS

CARLA QUARTE
Notary Public, Arizona
Maricopa County
My Comm. Expires Mar 29, 2013

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K.C. DAELLER #121 AGENCY LAKE HAVASU CITY POLICE DEPT.

DATE 072910 TIME 1109

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD MAIN JAIL

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution ___ AC. into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C + 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.099 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K.C. Daeller #121

K.C. DAELLER #121
DETECTIVE

DPS Form Exh G-2 (Rev 05-01)

[illegible]

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K.C. Dreller #121 AGENCY Lake Havasu City Police Dept

DATE 072910 TIME 1117

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD Main Jail

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
1. Clock time check.
2. Date check.

OPERATIONAL TESTS
1. Alcohol-free subject test result 0. 000 AC.
2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE K.C. Dreller #121

K.C. Dreller #121
Detective

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
07/29/2010 11:17:20

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

OAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SOAP,FREE
DOB: 07/29/2000
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:18:43
Diagnostic Test	Pass	11:19:17
Air Blank	0.000	11:19:44
0.100 Cal Check	0.099	11:20:05
Air Blank	0.000	11:20:34
Subject Test	0.000	11:20:57
Air Blank	0.000	11:21:24
Wait	ABT*	
Air Blank	0.000	11:22:06

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
07/29/2010 11:17:20

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

OAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SOAP,FREE
DOB: 07/29/2000
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:18:43
Diagnostic Test	Pass	11:19:17
Air Blank	0.000	11:19:44
0.100 Cal Check	0.099	11:20:05
Air Blank	0.000	11:20:34
Subject Test	0.000	11:20:57
Air Blank	0.000	11:21:24
Wait	ABT*	
Air Blank	0.000	11:22:06

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8.05.48
07/29/2000 11:17:20

Standard Lot#: 712322
Last Changed By: K. DREILER #12

OAS: K. DREILER #12
LAKE HAVASU CITY PD

Operator: K. DREILER #121
LAKE HAVASU CITY PD

Subject: SOAP,FREE
DOB: 07/29/2000
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:18:43
Diagnostic Test	Pass	11:19:17
Air Blank	0.000	11:19:44
0.100 Cal Check	0.099	11:20:05
Air Blank	0.000	11:20:34
Subject Test	0.000	11:20:57
Air Blank	0.000	11:21:24
Wait	ABT*	
Air Blank	0.000	11:22:06

*Sequence Aborted

Not a Successfully Completed Test Sequence 121

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
07/29/2010 11:23:26

Standard Lot#: 712322
Last Changed By: K. DRELLER #.21

OAS: K. DRELLER #.21
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SOAP,BART
DOB: 07/29/2000
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	TPS*	11:24:48
Air Blank	0.000	11:25:16

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
07/29/2010 11:23:26

Standard Lot#: 712322
Last Changed By: K. DRELLER #12.

OAS: K. DRELLER #12.
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SOAP,DAWT
DOB: 07/29/2000
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	11:24:48
Air Blank	0.000	11:25:16

*Improper Sample

Not a Successfully Completed Test Sequence

KCD 121

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
07/29/2010 11:23:26

Standard Lot#: 712322
Last Changed By: K. DREILER 4121

OAS: K. DREILER 4121
LAKE HAVASU CITY PD

Operator: K. DREILER 4121
LAKE HAVASU CITY PD

Subject: SOAP,BAWT
DOB: 07/29/2000
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	11:24:48
Air Blank	0.000	11:25:16

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
07/29/2010 11:23:07

Standard Data: 712022
Test Changed By: K. DRELLER #121

OAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SOAP,RFI
DOB: 07/29/2000
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	11:30:31
Air Blank	0.000	11:30:59

*RFI Detect

Not a Successfully Completed Test Sequence

121

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
07/29/2010 11:28:07

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

OAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SOAP,RFI
DOB: 07/29/2000
Sex: M Weight: 121

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	11:30:31
Air Blank	0.000	11:30:39

*RFI Detect

Not a Successfully Completed Test Sequence

[illegible]
Location: [illegible] MAIN [illegible]
Serial Number: [illegible]-001029
Core Version: [illegible]
07/29/2010 [illegible]

Standard Lot #: [illegible]
Last Changed By: K. [illegible] [illegible]

[illegible] [illegible]
[illegible] [illegible]

Operator: K. [illegible] [illegible]
[illegible] [illegible]

Subject: SCAP,RFI
DOB: 07/29/2000
Sex: M Weight: 121

1: Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	11:30:[illegible]
Air Blank	0.000	11:30:59

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 5000
Location: [illegible] MAIN [illegible]
Serial Number: 50-00[illegible]
Core Version: 8105.[illegible]8
07/29/2010 11:42:[illegible]

Standard Lot#: [illegible]
Last Changed By: K. DRELLER #121

GAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #12
LAKE HAVASU CITY PD

Subject: SQA2,DEPSAMP
DOB: 07/29/2000
Sex: M Weight: [illegible]

15 Minute Deprivation Period? [illegible]

Test	g/210L	Time
Air Blank	0.000	11:33:47
Diagnostic Test	Pass	11:34:21
Air Blank	0.000	11:34:48
0.100 Cal Check	0.10[illegible]	11:35:09
Air Blank	0.000	11:35:39
Subject Test	0.153*	11:38:46
Air Blank	0.000	11:39:18
Air Blank	0.000	11:39:46
Subject Test	XXX**	11:40:10
Air Blank	0.000	11:40:45

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 2000
Location: LACPD MAIN LAI
Serial Number: 80-001029
Core Version: 8.05.48
07/29/2010 [illegible]:52:42

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

OAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

Subject: SOAP,DEFSAMP
DOB: 07/29/2000
Sex: [illegible] Weight: 121

15 Minute Deprivation Period Yes

Test	g/210L	Time
Air Blank	0.000	11:33:47
Diagnostic Test	Pass	11:34:21
Air Blank	0.000	11:34:48
0.100 Cal Check	0.10	11:35:09
Air Blank	0.000	11:35:32
Subject Test	0.153*	11:38:46
Air Blank	0.000	11:39:18
Air Blank	0.000	11:39:46
Subject Test	XXX**	11:40:10
Air Blank	0.000	11:40:45

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 4000
[illegible]
Serial Number: [illegible]
Code Version: [illegible]
[illegible]

Standard Lot#: [illegible]
[illegible]

OAC: [illegible]
LAKE HAVASU [illegible]

Operator: [illegible]
LAKE HAVASU [illegible]

Subject: SOAP,[illegible]
DOB: 07/29/2000
Sex: [illegible] Weight: [illegible]

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:33:17
Diagnostic Test	Pass	11:34:21
Air Blank	0.000	11:34:48
0.100 Cal Check	0.101	11:35:09
Air Blank	0.000	11:35:39
Subject Test	0.153*	11:38:46
Air Blank	0.000	11:39:14
Air Blank	0.000	11:39:46
Subject Test	XXX**	11:40:10
Air Blank	0.000	11:40:45

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K.C. Dreller #121 AGENCY Lake Havasu City Police Dept

DATE 072910 TIME 1143

INTOXILYZER SERIAL # 80:001029 LOCATION LHCPD Main Jail

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C + 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed
5. Calibration check completed. Test results 0.099 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K.C. Dreller #121

K.C. Dreller #121
Detective

DPS Form Exh G-2 (Rev 05-01)