POLICE DEPARTMENT
**LAKE HAVASU CITY**
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Karl C. E. Dreller is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are a full, true and correct copy of the original.

31 DAY SIGNATURE 08132010

SN# 80-003356

STATE OF ARIZONA )
) ss
COUNTY OF MOHAVE )

Subscribed and sworn to before me this 24 day of SEPTEMBER, 2010 by CARLA DUARTE.

NOTARY PUBLIC

9-24-10
DATE

Commission expires: 3/29/2013

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2013

Area Code 928

Administration ..... 855-4884
Business Office ..... 855-1171
Investigations ..... 855-5775
Patrol ..... 855-0515

Fax #'s
Administration ..... 680-5430
Business Office ..... 680-5431
Investigations/Patrol ..... 680-5432

Emergency ..... 9-1-1
Non-Emergency Dispatch/ ..... 855-4111
TDD 855-4114
Havasu Silent Witness ..... 854-TIPS

E-mail: police@ci.lake-havasu-city.az.us

My Comm. Expires Mar 29, 2013

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST K.C. DREELER #121 AGENCY LAKE HAVASU CITY POLICE DEPT

DATE 081310 TIME 0830

INTOXILYZER SERIAL # 80-003356 LOCATION SIGN 8 LB CHANNEL

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test After selection is made press ENTER
4. Air blank completed.
5. Calibration check completed. Test results 0.101 AC
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE K.C. Dreeler #121

K.C. DREELER #121
DETECTIVE

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: SIGN 5 LB CHANNEL
Serial Number: 50 003356
Core Version: 8105.48
08/13/2010 08:30:52

---

Standard Lot#: 712323
Last Changed By: K. DRELLER #121

QAS: K. DRELLER #121
LAKE HAVASU CITY PD

---

Operator: K. DRELLER #121
LAKE HAVASU CITY PD

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 08:31:31 |
| 0.100 Cal Check | 0.101 | 08:31:52 |
| Air Blank | 0.000 | 08:32:21 |

K.C. [signature] 121
Preventative Maintenance
Performed -- Timer Reset