POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 18th day of August, 2015 by Josh Sondrol

Shawna McKnight
NOTARY PUBLIC

Aug 18, 2015
DATE

Commission expires: May 4, 2017

SHAWNA MCKNIGHT
Notary Public - Arizona
Mohave County
My Comm. Expires May 4, 2017

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	TDD	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

JS187

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LHC PD

DATE 8/18/15 TIME 0148

INTOXILYZER SERIAL # 80-003355 LOCATION CB CHANNEL BOOKING

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(✓) 4. Air blank completed.
(✓) 5. Calibration check completed. Test results 0. 099 AC.
(✓) 6. Air blank completed.
(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE JSondrol 187

DPS Form Exh G-2 (Rev 05-01)

*THIS INTOX JUST CAME BACK FROM DPS FOR QUARTERLY PM
* NOT IN SERVICE

JS187

INTOXILYZER 8000
Location: LB CHANNEL BOOKING
Serial Number: 80-003355
Core Version: 8105.48
08/18/2015 01:48:34

Standard Lot#: 203141003
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	01:49:05
0.100 Cal Check	0.099	01:49:26
Air Blank	0.000	01:49:55

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LNC PD

DATE 8/18/15 TIME 0152

INTOXILYZER SERIAL # 80-003355 LOCATION LB CHANNEL BO

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

*NOT IN SERVICE

SIGNATURE J Sondrol 187

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL BOOKING
Serial Number: 80-003355
Core Version: 8105.48
08/18/2015 01:56:33

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:57:20
Diagnostic Test	Pass	01:57:57
Air Blank	0.000	01:58:25
0.100 Cal Check	0.100	01:58:45
Air Blank	0.000	01:59:15
Subject Test	0.208*	02:02:22
Air Blank	0.000	02:02:56
Air Blank	0.000	02:03:23
Subject Test	XXX**	02:05:07
Air Blank	RFI***	02:05:21

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

ILYZER 8000
ion: LB CHANNEL BOOKING
l Number: 80-003355
Version: 8105.48
/20.5 01:52:09

ard Lot#: 20314100A3
Changed By: J. SONDROL #187

J. SONDROL #187
HAVASU CITY PD

tor: J. SONDROL #187
HAVASU CITY PD

ct: SQAP,TEST
01/01/1950
1 Weight: 13

nute Deprivation Period? Yes

	g/210L	Time
ank	0.000	01:52:58
stic Test	Pass	01:53:35
ank	0.000	01:54:03
Cal Check	0.098	01:54:24
ank	0.000	01:54:53
t Test	0.000	01:55:14
ank	ABT*	01:55:21
ank	IPS**	01:55:29

nce Aborted
oper Sample

t a Successfully
Completed
Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LHC PD

DATE 8/18/15 TIME 0209

INTOXILYZER SERIAL # 80-003355 LOCATION LB CHANNEL BOOKING

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

J 187

INTOXILYZER 8000
Location: LB CHANNEL BOOKING
Serial Number: 80-003355
Core Version: 8105.48
08/18/2015 02:09:07

Standard Lot#: 20314100H3
Last Changed By: J. SONDROL #187

ORS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	02:09:36
0.100 Cal Check	0.099	02:09:57
Air Blank	0.000	02:10:26

Preventative Maintenance
Performed -- Timer Reset