POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 30 **day of** AUGUST.

2013 **by** CARLA DUARTE

NOTARY PUBLIC

8/30/13

DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2017

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	*TDD*	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LHCRD

DATE 08-24-2013 TIME 1415 HRS

INTOXILYZER SERIAL # 80-003356 LOCATION LHC MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.098 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
08/24/2013 14:15:10

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	14:15:57
0.100 Cal Check	0.098	14:16:18
Air Blank	0.000	14:16:47

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LHCPD

DATE 08-24-2013 TIME 1418

INTOXILYZER SERIAL # 80-003356 LOCATION LHC MAIN JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 98 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
08/24/2013 14:37:18

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1959
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:38:12
Diagnostic Test	Pass	14:38:47
Air Blank	0.000	14:39:14
0.100 Cal Check	0.098	14:39:35
Air Blank	0.000	14:40:04
Subject Test	XXX*	14:40:25
Air Blank	RFI**	14:40:42

*Invalid Sample
Begin new deprivation period
**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
08/24/2013 14:37:18

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1959
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:38:12
Diagnostic Test	Pass	14:38:47
Air Blank	0.000	14:39:14
0.100 Cal Check	0.098	14:39:35
Air Blank	0.000	14:40:04
Subject Test	XXX*	14:40:25
Air Blank	RFI**	14:40:42

*Invalid Sample
Begin new deprivation period
**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
08/24/2013 14:37:18

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1959
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:38:12
Diagnostic Test	Pass	14:38:47
Air Blank	0.000	14:39:14
0.100 Cal Check	0.098	14:39:35
Air Blank	0.000	14:40:04
Subject Test	XXX*	14:40:25
Air Blank	RFI**	14:40:42

*Invalid Sample
Begin new deprivation period
**RFI Detect

Not a Successfully Completed Test Sequence

CT
219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
08/24/2013 14:23:47

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1959
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:24:41
Diagnostic Test	Pass	14:25:16
Air Blank	0.000	14:25:43
0.100 Cal Check	0.098	14:26:04
Air Blank	0.000	14:26:34
Subject Test	0.028	14:26:57
Air Blank	0.000	14:27:29
Wait		
Air Blank	0.000	14:32:25
Subject Test	0.046*	14:35:32
Air Blank	0.000	14:36:04

*Deficient Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
08/24/2013 14:23:47

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1959
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:24:41
Diagnostic Test	Pass	14:25:16
Air Blank	0.000	14:25:43
0.100 Cal Check	0.098	14:26:04
Air Blank	0.000	14:26:34
Subject Test	0.028	14:26:57
Air Blank	0.000	14:27:29
Wait		
Air Blank	0.000	14:32:25
Subject Test	0.046*	14:35:32
Air Blank	0.000	14:36:04

*Deficient Sample

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
08/24/2013 14:23:47

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1959
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:24:41
Diagnostic Test	Pass	14:25:16
Air Blank	0.000	14:25:43
0.100 Cal Check	0.098	14:26:04
Air Blank	0.000	14:26:34
Subject Test	0.028	14:26:57
Air Blank	0.000	14:27:29
Wait		
Air Blank	0.000	14:32:25
Subject Test	0.046*	14:35:32
Air Blank	0.000	14:36:04

*Deficient Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
08/24/2013 14:18:36

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1959
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:19:34
Diagnostic Test	Pass	14:20:09
Air Blank	0.000	14:20:36
0.100 Cal Check	0.099	14:20:57
Air Blank	0.000	14:21:27
Subject Test	0.000	14:21:44
Air Blank	ABT*	14:21:50
Air Blank	IPS**	14:21:58

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
08/24/2013 14:18:36

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1959
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:19:34
Diagnostic Test	Pass	14:20:09
Air Blank	0.000	14:20:36
0.100 Cal Check	0.099	14:20:57
Air Blank	0.000	14:21:27
Subject Test	0.000	14:21:44
Air Blank	ABT*	14:21:50
Air Blank	IPS**	14:21:58

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
08/24/2013 14:18:36

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1959
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:19:34
Diagnostic Test	Pass	14:20:09
Air Blank	0.000	14:20:36
0.100 Cal Check	0.099	14:20:57
Air Blank	0.000	14:21:27
Subject Test	0.000	14:21:44
Air Blank	ABT*	14:21:50
Air Blank	IPS**	14:21:58

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LHCPD

DATE 08-24-2013 TIME _____

INTOXILYZER SERIAL # 80-003356 LOCATION LHC MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.99 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
08/24/2013 14:43:38

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	14:44:09
0.100 Cal Check	0.099	14:44:29
Air Blank	0.000	14:44:59

Preventative Maintenance
Performed -- Timer Reset