POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

J Sondrol 187
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 10 day of SEPTEMBER, 2012 by CARLA DUARTE

Carla Duarte
NOTARY PUBLIC

9-10-12
DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2013

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	TDD	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL #187 AGENCY LAKE HAVASU CITY PD

DATE 9-10-12 TIME 1852

INTOXILYZER SERIAL # 80-003355 LOCATION LB CHANNEL SIGN 8

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.098 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

J/187

INTOXILYZER 5000
Location: LS CHANNEL SIGN 8
Serial Number: 80 003355
Core Version: 8.06.45
09/10/2012 18:52:56

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	18:53:31
0.100 Cal Check	0.098	18:54:52
Air Blank	0.000	18:54:21

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J SANDROL #187 AGENCY LAKE HAVASU CITY PD

DATE 9-10-12 TIME 1856

INTOXILYZER SERIAL # 80-003355 LOCATION LE CHANNEL SIGN 8

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC. Result 0.097 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE J Sandrol 187

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
09/10/2012 18:56:10

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:57:00
Diagnostic Test	Pass	18:57:35
Air Blank	0.000	18:58:02
0.100 Cal Check	0.098	18:58:23
Air Blank	0.000	18:58:53
Subject Test	0.000	18:59:12
Air Blank	ABT*	18:59:19
Air Blank	IPS**	18:59:26

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

9/187

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
09/10/2012 19:00:31

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:01:22
Diagnostic Test	Pass	19:01:57
Air Blank	0.000	19:02:24
0.100 Cal Check	0.097	19:02:45
Air Blank	0.000	19:03:15
Subject Test	0.287*	19:06:22
Air Blank	0.000	19:06:55
Air Blank	0.000	19:07:22
Subject Test	XXX**	19:07:51
Air Blank	RFI***	19:08:04

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

48181

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDREL #187 AGENCY LAKE HAVASU CITY PD

DATE 9-10-12 TIME 1909

INTOXILYZER SERIAL # 80-003355 LOCATION US CHANNEL SIGN 8

(✓) 1 Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ ____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.096 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrel 187

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LE CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 5105.46
09/10/2012 19:09:41

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	19:10:11
0.100 Cal Check	0.096	19:10:32
Air Blank	0.000	19:11:02

Preventative Maintenance
Performed -- Timer Reset