POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 22 day of Sept.
2013 by CHUCK TRESCHER

Shawna McKnight
NOTARY PUBLIC

9/22/2013
DATE

Commission expires: MAY 4, 2017

SHAWNA MCKNIGHT
Notary Public - Arizona
Mohave County
My Comm. Expires May 4, 2017

Area Code 928

		Fax #'s			
Administration	855-4884	Administration	680-5430	Emergency	9-1-1
Business Office	855-1171	Business Office	680-5431	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Investigations/Patrol	680-5432	*TDD*	855-4114
Patrol	855-0515			Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

CT
219

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST C. TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 09-22-2013 TIME 1523

INTOXILYZER SERIAL # 80-001029 LOCATION BRIDGEWATER CHANNEL SIGN 8

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(✓) 4. Air blank completed.
(✓) 5. Calibration check completed. Test results 0.098 AC.
(✓) 6. Air blank completed.
(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

C1
219

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
09/22/2013 15:23:32

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	15:24:02
0.100 Cal Check	0.098	15:24:23
Air Blank	0.000	15:24:52

Preventative Maintenance
Performed -- Timer Reset

C1
219

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST C. TRESCHER AGENCY LAKE HAVASU ~~CITY~~ POLICE

DATE 09-22-2013 TIME 1532

INTOXILYZER SERIAL # 80-001029 LOCATION BRIDGEWATER CHANNEL SIGN 8

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
09/22/2013 15:28:19

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:29:11
Diagnostic Test	Pass	15:29:46
Air Blank	0.000	15:30:13
0.100 Cal Check	0.097	15:30:34
Air Blank	0.000	15:31:03
Subject Test	0.000	15:31:22
Air Blank	ABT*	15:31:32
Air Blank	IPS**	15:31:40

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

a
219

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
09/22/2013 15:28:19

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:29:11
Diagnostic Test	Pass	15:29:46
Air Blank	0.000	15:30:13
0.100 Cal Check	0.097	15:30:34
Air Blank	0.000	15:31:03
Subject Test	0.000	15:31:22
Air Blank	ABT*	15:31:32
Air Blank	IPS**	15:31:40

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
09/22/2013 15:28:19

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:29:11
Diagnostic Test	Pass	15:29:46
Air Blank	0.000	15:30:13
0.100 Cal Check	0.097	15:30:34
Air Blank	0.000	15:31:03
Subject Test	0.000	15:31:22
Air Blank	ABT*	15:31:32
Air Blank	IPS**	15:31:40

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
09/22/2013 15:32:59

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:33:50
Diagnostic Test	Pass	15:34:25
Air Blank	0.000	15:34:53
0.100 Cal Check	0.098	15:35:14
Air Blank	0.000	15:35:43
Subject Test	0.072*	15:38:50
Air Blank	0.000	15:39:22
Air Blank	0.000	15:39:49
Subject Test	XXX**	15:40:10
Air Blank	RFI***	15:40:18

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

21c

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
09/22/2013 15:32:59

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:33:50
Diagnostic Test	Pass	15:34:25
Air Blank	0.000	15:34:53
0.100 Cal Check	0.098	15:35:14
Air Blank	0.000	15:35:43
Subject Test	0.072*	15:38:50
Air Blank	0.000	15:39:22
Air Blank	0.000	15:39:49
Subject Test	XXX**	15:40:10
Air Blank	RFI***	15:40:18

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
09/22/2013 15:32:59

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:33:50
Diagnostic Test	Pass	15:34:25
Air Blank	0.000	15:34:53
0.100 Cal Check	0.098	15:35:14
Air Blank	0.000	15:35:43
Subject Test	0.072*	15:38:50
Air Blank	0.000	15:39:22
Air Blank	0.000	15:39:49
Subject Test	XXX**	15:40:10
Air Blank	RFI***	15:40:18

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST C. TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 09-22-2013 TIME 1541

INTOXILYZER SERIAL # 80-001029 LOCATION BRIDGEWATER CHANNEL SIGN 8

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(✓) 4. Air blank completed.
(✓) 5. Calibration check completed. Test results 0.097 AC.
(✓) 6. Air blank completed.
(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

219

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-001029
Core Version: 8105.48
09/22/2013 15:41:41

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	15:42:11
0.100 Cal Check	0.097	15:42:31
Air Blank	0.000	15:43:01

Preventative Maintenance
Performed -- Timer Reset