POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 22 day of Sept.
2013 by CHUCK TRESCHER

Shawna McKnight
NOTARY PUBLIC

9/22/2013
DATE

Commission expires: May 1, 2017

SHAWNA MCKNIGHT
Notary Public - Arizona
Mohave County
My Comm. Expires May 4, 2017

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	*TDD*	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST C. TRESCHER AGENCY LAKE HAVASU POLICE

DATE 09-22-2013 TIME 1314 HRS

INTOXILYZER SERIAL # 80-003356 LOCATION LHCPD JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.098 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
09/22/2013 13:14:04

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	13:14:50
0.100 Cal Check	0.098	13:15:11
Air Blank	0.000	13:15:40

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST C TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 09-22-2013 TIME 1333 HRS

INTOXILYZER SERIAL # 80-003356 LOCATION LHCPD JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE

DPS Form Exh G-4 (Rev 05-01)

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
09/22/2013 13:23:24

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:24:42
Diagnostic Test	Pass	13:25:17
Air Blank	0.000	13:25:44
0.100 Cal Check	0.099	13:26:05
Air Blank	0.000	13:26:35
Subject Test	0.000	13:26:54
Air Blank	0.000	13:27:21
Wait		
Air Blank	ABT*	13:32:02
Air Blank	IPS**	13:32:10

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

C1
219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
09/22/2013 13:23:24

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:24:42
Diagnostic Test	Pass	13:25:17
Air Blank	0.000	13:25:44
0.100 Cal Check	0.099	13:26:05
Air Blank	0.000	13:26:35
Subject Test	0.000	13:26:54
Air Blank	0.000	13:27:21
Wait		
Air Blank	ABT*	13:32:02
Air Blank	IPS**	13:32:10

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
09/22/2013 13:23:24

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:24:42
Diagnostic Test	Pass	13:25:17
Air Blank	0.000	13:25:44
0.100 Cal Check	0.099	13:26:05
Air Blank	0.000	13:26:35
Subject Test	0.000	13:26:54
Air Blank	0.000	13:27:21
Wait		
Air Blank	ABT*	13:32:02
Air Blank	IPS**	13:32:10

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
09/22/2013 13:33:23

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:34:18
Diagnostic Test	Pass	13:34:52
Air Blank	0.000	13:35:20
0.100 Cal Check	0.098	13:35:41
Air Blank	0.000	13:36:10
Subject Test	0.061*	13:39:18
Air Blank	0.000	13:39:49
Air Blank	0.000	13:40:16
Subject Test	XXX**	13:40:35
Air Blank	RFI***	13:40:42

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
09/22/2013 13:33:23

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:34:18
Diagnostic Test	Pass	13:34:52
Air Blank	0.000	13:35:20
0.100 Cal Check	0.098	13:35:41
Air Blank	0.000	13:36:10
Subject Test	0.061*	13:39:18
Air Blank	0.000	13:39:49
Air Blank	0.000	13:40:16
Subject Test	XXX**	13:40:35
Air Blank	RFI***	13:40:42

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
09/22/2013 13:33:23

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:34:18
Diagnostic Test	Pass	13:34:52
Air Blank	0.000	13:35:20
0.100 Cal Check	0.098	13:35:41
Air Blank	0.000	13:36:10
Subject Test	0.061*	13:39:18
Air Blank	0.000	13:39:49
Air Blank	0.000	13:40:16
Subject Test	XXX**	13:40:35
Air Blank	RFI***	13:40:42

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST C. TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 09-22-13 TIME 1343 HRS

INTOXILYZER SERIAL # 80-003356 LOCATION LNCPD JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

DPS Form Exh G-2 (Rev 05-01)

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
09/22/2013 13:42:49

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	AMB*	13:43:19
Air Blank	0.000	13:43:48

*Ambient Fail

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
09/22/2013 13:43:54

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	13:44:23
0.100 Cal Check	0.097	13:44:44
Air Blank	0.000	13:45:14

Preventative Maintenance
Performed -- Timer Reset