POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Earl Wilson is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 23 day of OCTOBER, 2012 by CARLA DUARTE

NOTARY PUBLIC

10-23-12
DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2013

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
TDD 855-4114
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Earl Wilson AGENCY Lake Havasu City P.D.

DATE 10/10/12 TIME 12:04:37

INTOXILYZER SERIAL # 80-001029 LOCATION Jail

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.098 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [signature]

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
10/10/2012 12:04:37

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	12:05:08
0.100 Cal Check	0.098	12:05:29
Air Blank	0.000	12:05:59

Preventative Maintenance
Performed -- Timer Reset

EW
226

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Earl Wilson 224 AGENCY Lake Havasu City P.D.

DATE 10/10/12 TIME 11:39:17

INTOXILYZER SERIAL # 80-001029 LOCATION Jail

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

RFI Detect @ 11:49:55 on Deficient Sample Test. Prior to Invalid Sample

Retest @ 11:52:49 Successful. No Error Either Test

SIGNATURE EW 224

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
10/10/2012 11:39:17

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/2010
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:40:18
Diagnostic Test	Pass	11:40:53
Air Blank	0.000	11:41:20
0.100 Cal Check	0.099	11:41:41
Air Blank	0.000	11:42:11
Subject Test	0.000	11:42:28
Air Blank	ABT*	11:42:35
Air Blank	IPS**	11:42:43

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

Etw

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
10/10/2012 11:43:55

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/2010
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:44:57
Diagnostic Test	Pass	11:45:32
Air Blank	0.000	11:46:00
0.100 Cal Check	0.099	11:46:21
Air Blank	0.000	11:46:50
Subject Test	RFI*	11:49:04
Air Blank	0.000	11:49:37

*RFI Detect

Not a Successfully Completed Test Sequence

EW 226

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
10/10/2012 11:52:49

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/2010
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:53:52
Diagnostic Test	Pass	11:54:27
Air Blank	0.000	11:54:54
0.100 Cal Check	0.099	11:55:15
Air Blank	0.000	11:55:45
Subject Test	0.271*	11:58:52
Air Blank	0.000	11:59:25
Air Blank	0.000	11:59:53
Subject Test	XXX**	12:00:11
Air Blank	RFI***	12:00:28

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EW 22

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Earl Wilson 226 AGENCY Lake Havasu City P.D.

DATE 10/10/12 TIME 11:35:52

INTOXILYZER SERIAL # 80-001029 LOCATION Jail

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Earl Wilson 226

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
10/10/2012 11:35:52

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	11:36:25
0.100 Cal Check	0.099	11:36:46
Air Blank	0.000	11:37:16

Preventative Maintenance
Performed -- Timer Reset

CAW 226