POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Earl Wilson is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 23 day of OCTOBER, 2012 by CARLA DUARTE

NOTARY PUBLIC

10-23-12
DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2013

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	TDD	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Earl Wilson 226 AGENCY Lake Havasu City P.D.

DATE 10/10/12 TIME 10:32:03

INTOXILYZER SERIAL # 80-003355 LOCATION Bridgewater Channel Sign #8

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100AC.

OR

Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.096 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Earl Wilson 226

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
10/10/2012 10:32:03

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	10:32:34
0.100 Cal Check	0.096	10:32:55
Air Blank	0.000	10:33:24

Preventative Maintenance
Performed -- Timer Reset

CW 226

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST EARL WILSON #226 AGENCY LAKE HAVASU CITY P.D.

DATE 10/10/12 TIME 10:39:05

INTOXILYZER SERIAL # 80-003355 LOCATION BRIDGEWATER CHANNEL SIGN #8

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC. Result 0.097 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

ON DEFICIENT SAMPLE TEST @ 10:44:10 RFI WAS DETECTED PRIOR TO INVALID SAMPLE.

RETEST @ 10:52:55 WAS SUCCESSFUL SEQUENCE. NOTE: NO ERRORS ON EITHER TEST.

SIGNATURE Earl Wilson 226

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
10/10/2012 10:39:05

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/2010
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:40:26
Diagnostic Test	Pass	10:41:01
Air Blank	0.000	10:41:28
0.100 Cal Check	0.097	10:41:49
Air Blank	0.000	10:42:19
Subject Test	0.000	10:42:35
Air Blank	ABT*	10:42:42
Air Blank	IPS**	10:42:51

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

EW 226

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
10/10/2012 10:44:10

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/2010
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:45:16
Diagnostic Test	Pass	10:45:51
Air Blank	0.000	10:46:19
0.100 Cal Check	0.097	10:46:40
Air Blank	0.000	10:47:09
Subject Test	0.288*	10:50:17
Air Blank	0.000	10:50:50
Air Blank	0.000	10:51:17
Subject Test	RFI**	10:51:35
Air Blank	0.000	10:52:09

*Deficient Sample
**RFI Detect

Not a Successfully Completed Test Sequence

EAW
226

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
10/10/2012 10:52:55

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/10/2010
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:54:03
Diagnostic Test	Pass	10:54:38
Air Blank	0.000	10:55:06
0.100 Cal Check	0.096	10:55:27
Air Blank	0.000	10:55:56
Subject Test	0.560*	10:59:04
Air Blank	0.000	10:59:41
Air Blank	0.000	11:00:09
Subject Test	XXX**	11:00:26
Air Blank	RFI***	11:00:33

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EW 226

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Earl Wilson 226 AGENCY Lake Havasu City P.D.

DATE 10/10/12 TIME 11:05:38

INTOXILYZER SERIAL # 88-003355 LOCATION Bridgewater Channel Sign #8

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.095 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Earl Wilson 226

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003359
Core Version: 8105.48
10/10/2012 11:05:38

Standard Lot#: 35710100A1
Last Changed By: E. WILSON #226

QAS: E. WILSON #226
LAKE HAVASU CITY PD

Operator: E. WILSON #226
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	11:06:09
0.100 Cal Check	0.095	11:06:30
Air Blank	0.000	11:07:00

Preventative Maintenance
Performed -- Timer Reset

OW 226