POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

JSondrol
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 5 day of NOVEMBER, 2012 by CARLA DUARTE

Carla Duarte
NOTARY PUBLIC

11-5-12
DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2013

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	TDD	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LAKE HAVASU CITY PD

DATE 11-4-12 TIME 1950

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

9/187

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
11/04/2012 19:50:47

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	19:51:19
0.100 Cal Check	0.099	19:51:39
Air Blank	0.000	19:52:09

Preventative Maintenance
Performed Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J. SONDROL 187 AGENCY LAKE HAVASU CITY PD

DATE 11-4-12 TIME 1955

INTOXILYZER SERIAL # 80-001029 LOCATION LHCPD MAIN JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0.000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0.100 AC. Result 0.099 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE J Sondrol 187

DPS Form Exh G-4 (Rev 05-01)

8187

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
11/04/2012 19:55:20

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 11/04/2010
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:56:49
Diagnostic Test	Pass	19:57:24
Air Blank	0.000	19:57:51
0.100 Cal Check	0.099	19:58:12
Air Blank	0.000	19:58:42
Subject Test	0.000	19:59:04
Air Blank	ABT*	19:59:11
Air Blank	IPS**	19:59:18

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
11/04/2012 19:55:20

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 11/04/2010
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:56:49
Diagnostic Test	Pass	19:57:24
Air Blank	0.000	19:57:51
0.100 Cal Check	0.099	19:58:12
Air Blank	0.000	19:58:42
Subject Test	0.000	19:59:04
Air Blank	ABT*	19:59:11
Air Blank	IPS**	19:59:18

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
11/04/2012 19:55:20

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 11/04/2010
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:56:49
Diagnostic Test	Pass	19:57:24
Air Blank	0.000	19:57:51
0.100 Cal Check	0.099	19:58:12
Air Blank	0.000	19:58:42
Subject Test	0.000	19:59:04
Air Blank	ABT*	19:59:11
Air Blank	IPS**	19:59:18

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

7187

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
11/04/2012 20:00:46

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 11/04/2010
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	20:01:57
Diagnostic Test	Pass	20:02:32
Air Blank	0.000	20:02:59
0.100 Cal Check	0.099	20:03:20
Air Blank	0.000	20:03:50
Subject Test	0.235*	20:06:57
Air Blank	0.000	20:07:30
Air Blank	0.000	20:07:57
Subject Test	XXX**	20:08:17
Air Blank	RFI***	20:08:24

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

9 187

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
11/04/2012 20:00:46

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 11/04/2010
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	20:01:57
Diagnostic Test	Pass	20:02:32
Air Blank	0.000	20:02:59
0.100 Cal Check	0.099	20:03:20
Air Blank	0.000	20:03:50
Subject Test	0.235*	20:06:57
Air Blank	0.000	20:07:30
Air Blank	0.000	20:07:57
Subject Test	XXX**	20:08:17
Air Blank	RFI***	20:08:24

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

4187

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
11/04/2012 20:00:46

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 11/04/2010
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	20:01:57
Diagnostic Test	Pass	20:02:32
Air Blank	0.000	20:02:59
0.100 Cal Check	0.099	20:03:20
Air Blank	0.000	20:03:50
Subject Test	0.235*	20:06:57
Air Blank	0.000	20:07:30
Air Blank	0.000	20:07:57
Subject Test	XXX**	20:08:17
Air Blank	RFI***	20:08:24

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LAKE HAVASU CITY PD

DATE 11-4-12 TIME 2010

INTOXILYZER SERIAL # 80-001029 LOCATION LHC PD MAIN JAIL

( ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

( ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

( ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

( ) 4. Air blank completed.

( ) 5. Calibration check completed. Test results 0.099 AC.

( ) 6. Air blank completed.

( ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

Q18.

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
11/04/2012 20:10:39

Standard Lot#: 22911100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	20:11:09
0.100 Cal Check	0.099	20:11:30
Air Blank	0.000	20:11:59

Preventative Maintenance
Performed -- Timer Reset