POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Karl C. E. Dreller is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are a full, true and correct copy of the original.

K. C. Dreller #121
SIGNATURE
MAINTENANCE RECORD
80-001029

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 7 day of FEBRUARY, 2011 by CARLA DUARTE.

Carla Duarte
NOTARY PUBLIC

2-7-11
DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2013

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

E-mail: police@ci.lake-havasu-city.az.us

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
TDD 855-4114
Havasu Silent Witness 854-TIPS

ARIZONA DEPARTMENT OF PUBLIC SAFETY

INTOXILYZER SERVICE REQUEST

QUALITY ASSURANCE SPECIALIST

DATE	INTOXILYZER SERIAL NUMBER
11/23/10.	80-001029

SUMBITTING AGENCY: LHCPD

QUALITY ASSURANCE SPECIALIST	CELL PHONE NUMBER	PHONE NUMBER

EQUIPMENT SENT:

☒ INTOXILYZER ☐ POWER CORD ☐ GAS CYLINDER ☐ OTHER _____

PROBLEM DESCRIPTION: *(Include observations such as processor errors or other displayed messages. Attach sample printcards, if available.)*

Low CAL Checks

LAB USE ONLY

CORRECTIVE ACTION

Replace Breathose + Source Block Check Valve
Verify Calibration + Function Check

TECHNICIAN'S NAME	BADGE / ID NO.	TITLE	DATE
[signature]	6101	Electech.	11-30-10

INTOXILYZER 8000
Location: LHCPD MAIN JAIL
Serial Number: 80-001029
Core Version: 8105.48
11/30/2010 08:17:29

Standard Lot#: 712322
Last Changed By: K. DRELLER #121

OAS: K. DRELLER #121
LAKE HAVASU CITY PD

Operator: J. SLOWINSKI #6101
AZ DPS

Subject: TWENTY FOUR,SAMPLE,T
DOB: 03/01/1965
Sex: M Weight: 165

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:18:13
Diagnostic Test	Pass	08:18:50
Air Blank	0.000	08:19:18
0.100 Cal Check	0.101	08:19:39
Air Blank	0.000	08:20:08
Subject Test	0.187	08:20:41
Air Blank	0.000	08:21:15
Wait		
Air Blank	0.000	08:26:09
Subject Test	0.187	08:26:27
Air Blank	0.000	08:27:01
0.100 Cal Check	0.101	08:27:22
Air Blank	0.000	08:27:51
Diagnostic Test	Pass	08:28:29

Successfully
Completed
Test Sequence