POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 17 day of December, 2015 by Charles Trescher

Carla Duarte
NOTARY PUBLIC

12-17-15
DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2017

Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
TDD 855-4114
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 12-17-15 TIME 1528 HRS

INTOXILYZER SERIAL # 80-001873 LOCATION LHCPD MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
12/17/2015 15:28:27

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	15:29:36
0.100 Cal Check	0.100	15:29:57
Air Blank	0.000	15:30:27

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 12-17-15 TIME 1531 HRS

INTOXILYZER SERIAL # 80-001873 LOCATION LHCPD MAIN JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
12/17/2015 15:31:57

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:32:54
Diagnostic Test	Pass	15:33:29
Air Blank	0.000	15:33:56
0.100 Cal Check	0.100	15:34:17
Air Blank	0.000	15:34:47
Subject Test	0.000	15:35:10
Air Blank	ABT*	15:35:21
Air Blank	IPS**	15:35:29

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
12/17/2015 15:31:57

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:32:54
Diagnostic Test	Pass	15:33:29
Air Blank	0.000	15:33:56
0.100 Cal Check	0.100	15:34:17
Air Blank	0.000	15:34:47
Subject Test	0.000	15:35:10
Air Blank	ABT*	15:35:21
Air Blank	IPS**	15:35:29

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
12/17/2015 15:31:57

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:32:54
Diagnostic Test	Pass	15:33:29
Air Blank	0.000	15:33:56
0.100 Cal Check	0.100	15:34:17
Air Blank	0.000	15:34:47
Subject Test	0.000	15:35:10
Air Blank	ABT*	15:35:21
Air Blank	IPS**	15:35:29

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
12/17/2015 15:36:43

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:37:34
Diagnostic Test	Pass	15:38:09
Air Blank	0.000	15:38:36
0.100 Cal Check	0.100	15:38:57
Air Blank	0.000	15:39:27
Subject Test	0.042*	15:42:34
Air Blank	0.000	15:43:05
Air Blank	0.000	15:43:33
Subject Test	XXX**	15:44:01
Air Blank	RFI***	15:44:12

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
12/17/2015 15:36:43

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:37:34
Diagnostic Test	Pass	15:38:09
Air Blank	0.000	15:38:36
0.100 Cal Check	0.100	15:38:57
Air Blank	0.000	15:39:27
Subject Test	0.042*	15:42:34
Air Blank	0.000	15:43:05
Air Blank	0.000	15:43:33
Subject Test	XXX**	15:44:01
Air Blank	RFI***	15:44:12

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
12/17/2015 15:36:43

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:37:34
Diagnostic Test	Pass	15:38:09
Air Blank	0.000	15:38:36
0.100 Cal Check	0.100	15:38:57
Air Blank	0.000	15:39:27
Subject Test	0.042*	15:42:34
Air Blank	0.000	15:43:05
Air Blank	0.000	15:43:33
Subject Test	XXX**	15:44:01
Air Blank	RFI***	15:44:12

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 12-17-15 TIME 1545 HRS

INTOXILYZER SERIAL # 80-001873 LOCATION LHCPD MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
12/17/2015 15:45:26

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	15:45:56
0.100 Cal Check	0.100	15:46:16
Air Blank	0.000	15:46:46

Preventative Maintenance
Performed -- Timer Reset