POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 5 day of February.

2015 by Josh Sondrol

NOTARY PUBLIC

2-5-15

DATE

Commission expires:

CARLA DUARTE
Notary Public - Arizona
Mohave County
My Comm. Expires Mar 29, 2017

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDRUL 187 AGENCY LHC PD

DATE 2/4/15 TIME 1211

INTOXILYZER SERIAL # 80-003356 LOCATION LHC MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrul 187

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
02/04/2015 12:11:02

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	12:11:42
0.100 Cal Check	0.097	12:12:03
Air Blank	0.000	12:12:32

Preventative Maintenance
Performed -- Timer Reset

1

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LHC PD

DATE 2/4/15 TIME 1212

INTOXILYZER SERIAL # 80-003356 LOCATION LHC MAIN JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE J Sond 187

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
02/04/2015 12:12:42

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:13:42
Diagnostic Test	Pass	12:14:17
Air Blank	0.000	12:14:44
0.100 Cal Check	0.097	12:15:05
Air Blank	0.000	12:15:34
Subject Test	0.000	12:16:07
Air Blank	ABT*	12:16:14
Air Blank	IPS**	12:16:22

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
02/04/2015 12:12:42

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:13:42
Diagnostic Test	Pass	12:14:17
Air Blank	0.000	12:14:44
0.100 Cal Check	0.097	12:15:05
Air Blank	0.000	12:15:34
Subject Test	0.000	12:16:07
Air Blank	ABT*	12:16:14
Air Blank	IPS**	12:16:22

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
02/04/2015 12:12:42

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:13:42
Diagnostic Test	Pass	12:14:17
Air Blank	0.000	12:14:44
0.100 Cal Check	0.097	12:15:05
Air Blank	0.000	12:15:34
Subject Test	0.000	12:16:07
Air Blank	ABT*	12:16:14
Air Blank	IPS**	12:16:22

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
02/04/2015 12:17:20

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:18:08
Diagnostic Test	Pass	12:18:43
Air Blank	0.000	12:19:10
0.100 Cal Check	0.097	12:19:31
Air Blank	0.000	12:20:01
Subject Test	0.359*	12:23:08
Air Blank	0.000	12:23:42
Air Blank	0.000	12:24:09
Subject Test	XXX**	12:24:36
Air Blank	RFI***	12:24:43

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
02/04/2015 12:17:20

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:18:08
Diagnostic Test	Pass	12:18:43
Air Blank	0.000	12:19:10
0.100 Cal Check	0.097	12:19:31
Air Blank	0.000	12:20:01
Subject Test	0.359*	12:23:08
Air Blank	0.000	12:23:42
Air Blank	0.000	12:24:09
Subject Test	XXX**	12:24:36
Air Blank	RFI***	12:24:43

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
02/04/2015 12:17:20

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:18:08
Diagnostic Test	Pass	12:18:43
Air Blank	0.000	12:19:10
0.100 Cal Check	0.097	12:19:31
Air Blank	0.000	12:20:01
Subject Test	0.359*	12:23:08
Air Blank	0.000	12:23:42
Air Blank	0.000	12:24:09
Subject Test	XXX**	12:24:36
Air Blank	RFI***	12:24:43

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDRCR 187 AGENCY LNC PD

DATE 2/4/15 TIME 1226

INTOXILYZER SERIAL # 80-003356 LOCATION LNC MAIN JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.096 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondr 187

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-003356
Core Version: 8105.48
02/04/2015 12:26:14

Standard Lot#: 01513100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	12:26:43
0.100 Cal Check	0.096	12:27:04
Air Blank	0.000	12:27:34

Preventative Maintenance
Performed -- Timer Reset