POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 16th day of January, 2016 by Charles TRESCHER.

NOTARY PUBLIC

Jan 16, 2016
DATE

Commission expires: Feb. 24, 2017

WANDA R. SKINNER
Notary Public - Arizona
Mohave County
My Comm. Expires Feb 24, 2017

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	*TDD*	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 16th day of January, 2016 by WANDA R. SKINNER

NOTARY PUBLIC

Jan. 16, 2016
DATE

Commission expires: Feb 24, 2017

WANDA R. SKINNER
Notary Public - Arizona
Mohave County
My Comm. Expires Feb 24, 2017

Area Code 928

Administration	855-4884	Fax #'s		Emergency	9-1-1
Business Office	855-1171	Administration	680-5430	Non-Emergency Dispatch/	855-4111
Investigations	855-5775	Business Office	680-5431	*TDD*	855-4114
Patrol	855-0515	Investigations/Patrol	680-5432	Havasu Silent Witness	854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCAER AGENCY LAKE HAVASU CITY POLICE

DATE 01-16-16 TIME 1219 HRS

INTOXILYZER SERIAL # 80-001873 LOCATION LHCPD JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
01/16/2016 12:19:04

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	12:19:39
0.100 Cal Check	0.100	12:19:59
Air Blank	0.000	12:20:29

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST CHARLES TRESCHER AGENCY LAKE HAVASU CITY POLICE

DATE 01-16-16 TIME 1226 HRS

INTOXILYZER SERIAL # 80-001873 LOCATION LHCPD JAIL

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
(✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS
(✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES ✓ NO ____

COMMENTS

SIGNATURE

DPS Form Exh G-4 (Rev 05-01)

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
01/16/2016 12:22:03

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 215

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:23:06
Diagnostic Test	Pass	12:23:41
Air Blank	0.000	12:24:08
0.100 Cal Check	0.099	12:24:29
Air Blank	0.000	12:24:59
Subject Test	0.000	12:25:20
Air Blank	ABT*	12:25:26
Air Blank	IPS**	12:25:34

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
01/16/2016 12:22:03

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 215

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:23:06
Diagnostic Test	Pass	12:23:41
Air Blank	0.000	12:24:08
0.100 Cal Check	0.099	12:24:29
Air Blank	0.000	12:24:59
Subject Test	0.000	12:25:20
Air Blank	ABT*	12:25:26
Air Blank	IPS**	12:25:34

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
01/16/2016 12:22:03

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 215

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:23:06
Diagnostic Test	Pass	12:23:41
Air Blank	0.000	12:24:08
0.100 Cal Check	0.099	12:24:29
Air Blank	0.000	12:24:59
Subject Test	0.000	12:25:20
Air Blank	ABT*	12:25:26
Air Blank	IPS**	12:25:34

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
01/16/2016 12:26:10

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 215

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:27:13
Diagnostic Test	Pass	12:27:48
Air Blank	0.000	12:28:16
0.100 Cal Check	0.100	12:28:36
Air Blank	0.000	12:29:06
Subject Test	0.019*	12:32:13
Air Blank	0.000	12:32:43
Air Blank	0.000	12:33:11
Subject Test	XXX**	12:33:32
Air Blank	RFI***	12:33:44

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

219

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
01/16/2016 12:26:10

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 215

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:27:13
Diagnostic Test	Pass	12:27:48
Air Blank	0.000	12:28:16
0.100 Cal Check	0.100	12:28:36
Air Blank	0.000	12:29:06
Subject Test	0.019*	12:32:13
Air Blank	0.000	12:32:43
Air Blank	0.000	12:33:11
Subject Test	XXX**	12:33:32
Air Blank	RFI***	12:33:44

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
01/16/2016 12:26:10

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 10/19/1978
Sex: M Weight: 215

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:27:13
Diagnostic Test	Pass	12:27:48
Air Blank	0.000	12:28:16
0.100 Cal Check	0.100	12:28:36
Air Blank	0.000	12:29:06
Subject Test	0.019*	12:32:13
Air Blank	0.000	12:32:43
Air Blank	0.000	12:33:11
Subject Test	XXX**	12:33:32
Air Blank	RFI***	12:33:44

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST CHARLES TRESCHOCK AGENCY LAKE HAVASU CITY POLICE

DATE 01-16-16 TIME 1234 HRS

INTOXILYZER SERIAL # 80-001873 LOCATION LHCPD JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.100 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
01/16/2016 12:34:57

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: C. TRESCHER #219
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	12:35:28
0.100 Cal Check	0.100	12:35:49
Air Blank	0.000	12:36:19

Preventative Maintenance
Performed -- Timer Reset